

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000719

GEBS Results

GEBS VALUE = 0.100

Lab #: 071904230 12/01/2016 02:10:51

0.100 02:13

0.099 02:18

Test Message: 0.16

Lab #: 071904233 12/14/2016 00:26:35

0.099 00:28

0.098 00:33

Test Message: 0.17

Lab #: 071904234 12/20/2016 23:41:22

0.100 23:44

0.099 23:48

Test Message: 0.15

Lab #: 071904235 12/24/2016 02:47:56

0.099 02:50

0.099 02:56

Test Message: 0.17

Lab #: 071904236 12/30/2016 20:17:08

0.099 20:19

0.096 20:26

Test Message: 0.13

Lab #: 071904237 12/31/2016 00:17:43

0.099 00:19

0.098 00:24

Test Message: 0.24

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

Lab #: 072703619 12/02/2016 23:15:36

0.099 23:17

0.097 23:24

Test Message: 0.19

Lab #: 072703620 12/04/2016 03:23:37

0.098 03:25

0.096 03:30

Test Message: 0.19

Lab #: 072703625 12/13/2016 22:30:36

*Refer to original breath strip
for accuracy checks of*

0.099 01:33

0.098 01:38

0.098 22:32

0.095 22:38

Test Message: 0.20

Lab #: 072703626 12/14/2016 03:41:00

0.098 03:43

0.094 03:47

Test Message: 0.27

Lab #: 072703627 12/15/2016 07:00:14

0.098 07:02

0.096 07:08

Test Message: 0.13

Lab #: 072703628 12/19/2016 03:02:46

0.098 03:05

0.098 03:10

Test Message: 0.15

Lab #: 072703629 12/21/2016 18:42:58

0.099 18:45

0.097 18:50

Test Message: 0.07

Lab #: 072703630 12/24/2016 03:13:24

0.098 03:16

0.096 03:22

Test Message: 0.14

Lab #: 072703631 12/24/2016 20:05:46

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

Lab #: 072703631 12/24/2016 20:05:46

0.098 20:09

0.097 20:14

Test Message: 0.31

Lab #: 072703632 12/24/2016 20:48:51

0.099 20:50

0.098 20:56

Test Message: 0.10

Lab #: 072703633 12/24/2016 21:06:36

0.098 21:08

0.098 21:13

Test Message: 0.13

SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000727
12/05/16 22:58

Lab ID# = 072703622
GEBS Value = 0.100 g/210L

Test	g/210L	Time
Air Blank	0.000	22:59
Diagnostic Test	Pass	22:59
Air Blank	0.000	23:00
GEBS Check	0.098	23:00
Air Blank	0.000	23:01
Subject Test	0.075	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Test	0.072	23:04
Air Blank	0.000	23:05
GEBS Check	0.097	23:05
Air Blank	0.000	23:06
Diagnostic Test	Pass	23:06
Air Blank	0.000	23:06

Successfully
Completed Test

Reported Result: .07% .07

Tronetta Administration Give:

Time First Observed: _____

Observed By: _____

Operator's Signature

[REDACTED]
SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000727
12/09/16 00:10

Lab ID# = 072703623
GEBS Value = 0.100 g/210L

[REDACTED]

Test	g/210L	Time
Air Blank	0.000	00:11
Diagnostic Test	Pass	00:11
Air Blank	0.000	00:12
GEBS Check	0.098	00:12
Air Blank	0.000	00:13
Subject Test	0.208	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:17

*Radio Frequency Interferent Detected

NOT a Successfully
Completed Test

Time First Observed: _____

Observed by: _____

Operator's Signature

[REDACTED]
SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000727
12/09/16 22:48

Lab ID# = 072703624
GEBS Value = 0.100 g/210L

[REDACTED]

Test	g/210L	Time
Air Blank	0.000	22:49
Diagnostic Test	Pass	22:49
Air Blank	0.000	22:50
GEBS Check	0.098	22:50
Air Blank	0.000	22:51
Subject Test	0.221	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Test	0.205	22:54
Air Blank	0.000	22:55
GEBS Check	0.095	22:55
Air Blank	0.000	22:56
Diagnostic Test	Pass	22:56
Air Blank	0.000	22:56

Successfully
Completed Test

Reported Result: .22/.20

Trombetta Admonishment Given

Time First Observed: [REDACTED]

Observed by: [REDACTED]

[REDACTED]
Operator's Signature

[REDACTED]
SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000727
12/19/16 03:02

Lab ID# = 072703628
GEBS Value = 0.100 g/210L
[REDACTED]

Test	g/210L	Time
Air Blank	0.000	03:04
Diagnostic Test	Pass	03:04
Air Blank	0.000	03:05
GEBS Check	0.098	03:05
Air Blank	0.000	03:06
Subject Test	0.161	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Test	0.155	03:09
Air Blank	0.000	03:10
GEBS Check	0.098	03:10
Air Blank	0.000	03:11
Diagnostic Test	Pass	03:11
Air Blank	0.000	03:11

Successfully
Completed Test

Reported Result: .16/.15

Trombetta Admonishment Given

Time First Observed: [REDACTED]

Observed by: [REDACTED]

[REDACTED]
Operator's Signature

SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006058
12/19/16 02:42

Lab ID# = 605801452
GEBS Value = 0.100 g/210L
[REDACTED]

Test	g/210L	Time
Air Blank	0.000	02:44
Diagnostic Test	Pass	02:44
Air Blank	0.000	02:44
GEBS Check	0.100	02:44
Air Blank	0.000	02:45
Subject Test	0.176	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Test	0.170	02:49
Air Blank	0.000	02:49
GEBS Check	INT*	02:50
Air Blank	PUR**	02:51

*Interferent Detected Choose Another Test
**Purge Fail

NOT a Successfully
Completed Test

Time First Observed: [REDACTED]

Observed by: [REDACTED]

[REDACTED]
Operator's Signature

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000728

GEBS Results

GEBS VALUE = 0.100

Lab #: 072802447 12/03/2016 23:02:16

0.099 23:04

0.099 23:09

Test Message: 0.08

Lab #: 072802448 12/04/2016 02:48:49

0.101 02:50

0.099 02:57

Test Message: 0.09

Lab #: 072802453 12/18/2016 22:38:49

0.100 22:41

0.099 22:46

Test Message: 0.11

Lab #: 072802454 12/27/2016 01:37:35

0.100 01:39

Test Message: Deficient Sample

Lab #: 072802455 12/30/2016 21:48:57

0.099 21:51

0.098 21:56

Test Message: 0.12

Lab #: 072802456 12/31/2016 00:17:37

0.099 00:19

0.096 00:25

Test Message: 0.20

[REDACTED]
SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000728
12/04/16 22:19

Lab ID# = 072802449
GEBS Value = 0.100 g/210L

[REDACTED]

Test	g/210L	Time
Air Blank	0.000	22:21
Diagnostic Test	Pass	22:21
Air Blank	0.000	22:21
GEBS Check	0.099	22:22
Air Blank	0.000	22:22
Subject Test	0.137	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Test	ABT*	22:26
Air Blank	0.000	22:27

*Sequence Aborted

NOT a Successfully
Completed Test

Time First Observed: _____

Observed by: _____

[REDACTED]

Operator's Signature

[REDACTED]
SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000728
12/05/16 02:12

Lab ID# = 072802450
GEBS Value = 0.100 g/210L

[REDACTED]

Test	g/210L	Time
Air Blank	0.000	02:13
Diagnostic Test	Pass	02:13
Air Blank	0.000	02:14
GEBS Check	0.100	02:14
Air Blank	0.000	02:14
Subject Test	0.100	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Test	0.101	02:19
Air Blank	0.000	02:20
GEBS Check	0.098	02:20
Air Blank	0.000	02:20
Diagnostic Test	Pass	02:20
Air Blank	0.000	02:21

Successfully
Completed Test

Reported Result: .10/.10

Trombetta Admonishment Given

Time First Observed: [REDACTED]

Observed by: [REDACTED]

[REDACTED]
Operator's Signature

INTOXILYZER 8000
Instrument Initialization
20:35 12/10/16

[REDACTED]

SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000728
12/11/16 01:51

Lab ID# = 072802452
GEBS Value = 0.100 g/210L

[REDACTED]

Test	g/210L	Time
Air Blank	0.000	01:52
Diagnostic Test	Pass	01:52
Air Blank	0.000	01:52
GEBS Check	0.099	01:53
Air Blank	0.000	01:53
Subject Test	0.150	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Test	0.153	01:58
Air Blank	0.000	01:58
GEBS Check	0.099	01:59
Air Blank	0.000	01:59
Diagnostic Test	Pass	01:59
Air Blank	0.000	02:00

Successfully
Completed Test

Reported Result: .15/.15

Tronbetta Admonishment Given

Time First Observed: [REDACTED]

Observed by: [REDACTED]

[REDACTED]
Operator's Signature

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500746 12/04/2016 01:37:38

0.099 01:40

0.098 01:45

Test Message: 0.13

Lab #: 605500747 12/16/2016 23:16:03

0.100 23:18

0.099 23:23

Test Message: 0.11

Lab #: 605500748 12/16/2016 23:57:45

0.099 23:59

0.098 00:04

Test Message: 0.10

Lab #: 605500749 12/17/2016 01:28:33

0.098 01:30

0.098 01:35

Test Message: 0.10

Lab #: 605500750 12/17/2016 02:17:27

0.099 02:19

0.099 02:24

Test Message: 0.10

Lab #: 605500751 12/17/2016 02:27:10

0.098 02:29

0.097 02:34

Test Message: 0.11

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801399 12/01/2016 04:29:45

0.100 04:32

0.100 04:39

Test Message: 0.12

Lab #: 605801400 12/01/2016 22:27:36

0.100 22:29

0.099 22:34

Test Message: 0.15

Lab #: 605801401 12/02/2016 18:27:28

0.099 18:29

0.099 18:34

Test Message: 0.29

Lab #: 605801402 12/03/2016 05:21:39

0.100 05:23

0.098 05:28

Test Message: 0.09

Lab #: 605801403 12/03/2016 21:31:44

0.099 21:34

0.099 21:39

Test Message: 0.27

Lab #: 605801404 12/04/2016 01:13:08

0.100 01:15

0.091 01:20

Test Message: 0.12

Lab #: 605801405 12/04/2016 03:20:39

0.100 03:23

Test Message: Sequence Aborted

Lab #: 605801406 12/04/2016 07:23:19

0.095 07:25

Test Message: Subject Test Refused

Lab #: 605801407 12/04/2016 07:31:38

0.100 07:33

0.098 07:40

Test Message: 0.18

Lab #: 605801408 12/04/2016 15:52:57

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801408 12/04/2016 15:52:57

0.099 15:55

0.098 16:00

Test Message: 0.21

Lab #: 605801409 12/04/2016 19:23:13

0.099 19:25

0.096 19:31

Test Message: 0.19

Lab #: 605801410 12/05/2016 22:47:50

0.099 22:50

0.097 22:57

Test Message: 0.22

Lab #: 605801411 12/06/2016 00:47:13

0.099 00:49

0.098 00:54

Test Message: 0.17

Lab #: 605801412 12/07/2016 03:36:05

0.099 03:38

0.097 03:43

Test Message: 0.23

Lab #: 605801413 12/07/2016 18:00:29

0.099 18:02

0.098 18:07

Test Message: 0.17

Lab #: 605801414 12/08/2016 01:10:56

0.100 01:13

0.097 01:18

Test Message: 0.23

Lab #: 605801415 12/08/2016 23:53:07

0.099 23:55

0.097 00:00

Test Message: 0.21

Lab #: 605801416 12/09/2016 00:22:16

0.099 00:24

Test Message: Sequence Aborted

Lab #: 605801417 12/09/2016 03:04:38

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801417 12/09/2016 03:04:38

0.100 03:07

Test Message: Radio Frequency Interferent Detected

Lab #: 605801419 12/09/2016 03:11:36

0.099 03:14

0.098 03:19

Test Message: 0.18

Lab #: 605801420 12/09/2016 22:50:15

0.098 22:53

0.097 22:58

Test Message: 0.23

Lab #: 605801422 12/10/2016 00:24:04

0.099 00:26

0.097 00:32

Test Message: 0.25

Lab #: 605801423 12/10/2016 01:55:23

0.100 01:57

0.098 02:02

Test Message: 0.17

Lab #: 605801424 12/10/2016 03:28:14

0.099 03:31

0.098 03:36

Test Message: 0.14

Lab #: 605801425 12/10/2016 03:55:23

0.099 04:00

0.098 04:05

Test Message: 0.16

Lab #: 605801426 12/10/2016 04:22:01

0.099 04:25

0.098 04:30

Test Message: 0.17

Lab #: 605801427 12/11/2016 04:03:24

0.099 04:05

0.098 04:10

Test Message: 0.15

Lab #: 605801428 12/11/2016 07:05:11

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801428 12/11/2016 07:05:11

0.099 07:08

0.096 07:16

Test Message: 0.19

Lab #: 605801429 12/11/2016 16:23:08

0.098 16:25

Test Message: No Sample Given

Lab #: 605801430 12/11/2016 16:51:10

0.099 16:53

Test Message: No Sample Given

Lab #: 605801431 12/11/2016 17:00:27

0.098 17:02

Test Message: Subject Test Refused

Lab #: 605801432 12/11/2016 22:22:59

0.098 22:26

0.097 22:31

Test Message: 0.18

Lab #: 605801433 12/12/2016 17:40:37

0.098 17:42

0.096 17:49

Test Message: 0.15

Lab #: 605801434 12/13/2016 02:43:47

Refer to original breath strip for accuracy checks JS

~~0.099 02:45 JS 1/24/17~~

~~0.098 02:51~~

Test Message: 0.10

Lab #: 605801435 12/13/2016 22:55:27

0.099 22:58

0.097 23:03

Test Message: 0.19

Lab #: 605801436 12/14/2016 03:29:39

Refer to original breath strip for accuracy checks JS

~~0.099 22:22 JS~~

~~0.098 03:32 1/24/17~~

Test Message: Radio Frequency Interferent Detected

Lab #: 605801437 12/14/2016 03:33:31

Refer to original breath strip for accuracy checks JS

~~0.100 02:14 JS~~

~~0.098 02:20 1/24/17~~

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801437 12/14/2016 03:33:31

Refer to original breath strip
for accuracy checks *gf*

~~0.098 03:35~~

gf
1/24/17

~~0.096 03:40~~

Test Message: 0.27

Lab #: 605801438 12/16/2016 00:34:05

Refer to original breath strip
for accuracy checks *gf*

~~0.100 03:33~~

gf
1/24/17

~~0.099 00:36~~

~~0.098 00:40~~

Test Message: 0.16

Lab #: 605801439 12/16/2016 20:01:52

Refer to original breath strip
for accuracy checks *gf*

~~0.099 01:53~~

gf
1/24/17

~~0.099 01:59~~

~~0.100 20:03~~

Test Message: Sequence Aborted

Lab #: 605801440 12/17/2016 03:15:14

0.099 03:17

0.096 03:25

Test Message: 0.20

Lab #: 605801441 12/17/2016 21:24:40

0.100 21:26

0.100 21:33

Test Message: 0.02

Lab #: 605801443 12/17/2016 21:53:33

0.100 21:55

0.097 22:01

Test Message: 0.23

Lab #: 605801444 12/17/2016 23:00:36

0.100 23:02

0.099 23:08

Test Message: 0.16

Lab #: 605801445 12/18/2016 00:26:33

0.100 00:28

0.097 00:34

Test Message: 0.15

Lab #: 605801446 12/18/2016 02:58:04

0.100 03:00

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801446 12/18/2016 02:58:04

0.098 03:06

Test Message: 0.14

Lab #: 605801447 12/18/2016 04:18:46

*Refer to original breath strip
for accuracy checks of*

0.100 04:20

of 1/24/17

0.098 04:25

Test Message: 0.20

Lab #: 605801448 12/18/2016 11:32:50

0.100 11:35

0.099 11:40

Test Message: 0.21

Lab #: 605801449 12/18/2016 17:14:47

0.101 17:16

0.098 17:24

Test Message: 0.21

Lab #: 605801450 12/18/2016 19:06:40

0.099 19:08

0.097 19:14

Test Message: 0.19

Lab #: 605801451 12/18/2016 20:14:57

0.100 20:17

0.100 20:22

Test Message: 0.08

Lab #: 605801452 12/19/2016 02:42:40

*Refer to original breath strip
for accuracy checks of*

0.100 02:44

of 1/24/17

0.099 02:50

Test Message: Interferent Detected Choose Another Test

Lab #: 605801453 12/19/2016 08:22:29

0.101 08:24

0.101 08:28

Test Message: 0.00

Lab #: 605801454 12/19/2016 11:31:28

0.101 11:33

0.101 11:37

Test Message: 0.00

Lab #: 605801455 12/20/2016 01:44:09

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801455 12/20/2016 01:44:09

0.100 01:46

Test Message: Interferent Detected Choose Another Test

Lab #: 605801456 12/20/2016 01:48:49

0.101 01:50

0.099 01:55

Test Message: 0.12

Lab #: 605801457 12/22/2016 20:40:05

0.098 20:43

Test Message: Subject Test Refused

Lab #: 605801458 12/22/2016 22:53:22

0.098 22:55

0.097 23:00

Test Message: 0.14

Lab #: 605801459 12/23/2016 04:21:35

0.099 04:50

0.094 04:55

Test Message: 0.15

Lab #: 605801460 12/23/2016 22:56:33

0.100 22:59

0.097 23:04

Test Message: 0.15

Lab #: 605801461 12/23/2016 23:58:26

0.100 00:00

0.099 00:06

Test Message: 0.07

Lab #: 605801463 12/24/2016 00:35:14

0.099 00:37

0.098 00:42

Test Message: 0.24

Lab #: 605801464 12/24/2016 09:52:41

0.100 09:55

0.097 10:01

Test Message: 0.20

Lab #: 605801465 12/24/2016 19:49:30

0.100 19:52

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801465 12/24/2016 19:49:30

0.098 19:58

Test Message: 0.17

Lab #: 605801466 12/24/2016 21:03:54

0.100 21:05

Test Message: Improper Sample

Lab #: 605801467 12/25/2016 04:42:46

0.100 04:45

0.099 04:50

Test Message: 0.20

Lab #: 605801468 12/25/2016 07:16:58

0.100 07:18

0.098 07:24

Test Message: 0.15

Lab #: 605801469 12/25/2016 14:37:59

0.100 14:42

Test Message: Deficient Sample

Lab #: 605801470 12/25/2016 18:57:15

0.099 18:59

0.098 19:04

Test Message: 0.14

Lab #: 605801471 12/25/2016 19:26:13

0.099 19:28

Test Message: No Sample Given

Lab #: 605801472 12/25/2016 22:36:56

0.100 22:39

0.097 22:46

Test Message: 0.18

Lab #: 605801473 12/25/2016 23:25:25

0.100 23:28

0.099 23:33

Test Message: 0.14

Lab #: 605801474 12/26/2016 01:01:43

0.100 01:03

0.097 01:09

Test Message: 0.13

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801475 12/26/2016 21:35:14

0.100 21:38

0.097 21:45

Test Message: 0.11

Lab #: 605801476 12/28/2016 16:22:22

0.099 16:24

0.096 16:31

Test Message: 0.21

Lab #: 605801477 12/28/2016 22:15:08

0.099 22:17

Test Message: Deficient Sample

Lab #: 605801478 12/28/2016 22:25:45

0.099 22:28

0.095 22:35

Test Message: 0.26

Lab #: 605801479 12/29/2016 23:09:04

0.099 23:11

0.099 23:16

Test Message: 0.15

Lab #: 605801480 12/30/2016 02:24:27

0.098 02:26

Test Message: Deficient Sample

Lab #: 605801481 12/30/2016 02:30:51

0.099 02:32

0.098 02:38

Test Message: 0.10

Lab #: 605801482 12/30/2016 02:52:55

0.099 02:55

Test Message: Radio Frequency Interferent Detected

Lab #: 605801483 12/30/2016 02:58:43

0.099 03:00

0.098 03:05

Test Message: 0.09

Lab #: 605801484 12/31/2016 01:50:25

0.099 01:52

0.097 01:57

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Test Message: 0.12

SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006058
12/16/16 00:34

Lab ID# = 605801438
GEBS Value = 0.100 g/210L

Test	g/210L	Time
Air Blank	0.000	00:35
Diagnostic Test	Pass	00:35
Air Blank	0.000	00:35
GEBS Check	0.099	00:36
Air Blank	0.000	00:36
Subject Test	0.161	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Test	0.161	00:40
Air Blank	0.000	00:40
GEBS Check	0.098	00:40
Air Blank	0.000	00:41
Diagnostic Test	Pass	00:41
Air Blank	0.000	00:42

Successfully
Completed Test

Trombetta Admonishment Given

Time First Observed: [REDACTED]

Observed by: [REDACTED]

SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006058
12/16/16 00:34

Lab ID# = 605801438
GEBS Value = 0.100 g/210L

Test	g/210L	Time
Air Blank	0.000	00:35
Diagnostic Test	Pass	00:35
Air Blank	0.000	00:35
GEBS Check	0.099	00:36
Air Blank	0.000	00:36
Subject Test	0.161	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Test	0.161	00:40
Air Blank	0.000	00:40
GEBS Check	0.098	00:40
Air Blank	0.000	00:41
Diagnostic Test	Pass	00:41
Air Blank	0.000	00:42

Successfully
Completed Test

Trombetta Admonishment Given

Time First Observed: [REDACTED]

Observed by: [REDACTED]

SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006058
12/16/16 00:34

Lab ID# = 605801438
GEBS Value = 0.100 g/210L

Test	g/210L	Time
Air Blank	0.000	00:35
Diagnostic Test	Pass	00:35
Air Blank	0.000	00:35
GEBS Check	0.099	00:36
Air Blank	0.000	00:36
Subject Test	0.161	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Test	0.161	00:40
Air Blank	0.000	00:40
GEBS Check	0.098	00:40
Air Blank	0.000	00:41
Diagnostic Test	Pass	00:41
Air Blank	0.000	00:42

Successfully
Completed Test

Trombetta Admonishment Given

Time First Observed: [REDACTED]

Observed by: [REDACTED]

[REDACTED]
SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000727
12/19/16 03:02

Lab ID# = 072703628
GEBS Value = 0.100 g/210L

[REDACTED]

Test	g/210L	Time
Air Blank	0.000	03:04
Diagnostic Test	Pass	03:04
Air Blank	0.000	03:05
GEBS Check	0.098	03:05
Air Blank	0.000	03:06
Subject Test	0.161	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Test	0.155	03:09
Air Blank	0.000	03:10
GEBS Check	0.098	03:10
Air Blank	0.000	03:11
Diagnostic Test	Pass	03:11
Air Blank	0.000	03:11

Successfully
Completed Test

Reported Result: .16/.15

Trombetta Admonishment Given

Time First Observed: [REDACTED]

Observed by: [REDACTED]

[REDACTED]
Operator's Signature

SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006058
12/19/16 02:42

Lab ID# = 605801452
GEBS Value = 0.100 g/210L

[REDACTED]

Test	g/210L	Time
Air Blank	0.000	02:44
Diagnostic Test	Pass	02:44
Air Blank	0.000	02:44
GEBS Check	0.100	02:44
Air Blank	0.000	02:45
Subject Test	0.176	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Test	0.170	02:49
Air Blank	0.000	02:49
GEBS Check	INT*	02:50
Air Blank	PUR**	02:51

*Interferent Detected Choose Another Test
**Purge Fail

NOT a Successfully
Completed Test

Time First Observed: [REDACTED]

Observed by: [REDACTED]

[REDACTED]
Operator's Signature

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902702 12/04/2016 00:07:25

0.099 00:10

0.098 00:15

Test Message: 0.13

Lab #: 605902703 12/04/2016 01:43:52

0.098 01:46

0.097 01:52

Test Message: 0.18

Lab #: 605902704 12/16/2016 23:33:58

0.093 23:36

0.094 23:41

Test Message: 0.08

Lab #: 605902705 12/17/2016 00:51:04

0.099 00:52

0.098 00:59

Test Message: 0.15

Lab #: 605902706 12/17/2016 01:17:37

0.099 01:19

0.098 01:24

Test Message: 0.13

Lab #: 605902707 12/17/2016 02:00:35

0.098 02:02

0.096 02:07

Test Message: 0.13

Lab #: 605902708 12/17/2016 02:11:41

0.098 02:13

0.097 02:19

Test Message: 0.13

Lab #: 605902709 12/17/2016 02:48:12

0.098 02:50

0.098 02:55

Test Message: 0.11

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000719

GEBS Results

GEBS VALUE = 0.100

Lab #: 071904222 11/01/2016 23:27:00

0.098 23:28

0.098 23:33

Test Message: 0.19

Lab #: 071904223 11/03/2016 22:01:37

0.098 22:03

0.098 22:08

Test Message: 0.15

Lab #: 071904224 11/05/2016 00:36:23

0.098 00:38

0.098 00:42

Test Message: 0.25

Lab #: 071904225 11/09/2016 01:09:18

0.099 01:11

0.098 01:15

Test Message: 0.12

Lab #: 071904226 11/13/2016 17:54:59

0.100 17:56

0.099 18:01

Test Message: 0.15

Lab #: 071904227 11/16/2016 01:08:58

0.099 01:10

0.097 01:18

Test Message: 0.16

Lab #: 071904228 11/24/2016 22:46:14

0.100 22:48

0.099 22:52

Test Message: 0.12

Lab #: 071904229 11/28/2016 23:36:03

0.099 23:39

0.099 23:43

Test Message: 0.15

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000726

GEBS Results

GEBS VALUE = 0.100

Lab #: 072600355 11/06/2016 17:33:55

0.098 17:37

0.098 17:42

Test Message: 0.06

Lab #: 072600356 11/11/2016 01:42:39

0.100 01:45

0.098 01:51

Test Message: 0.16

Lab #: 072600357 11/12/2016 01:41:02

0.100 01:43

0.099 01:48

Test Message: 0.16

Lab #: 072600358 11/13/2016 16:58:24

0.100 17:00

0.100 17:05

Test Message: 0.05

Lab #: 072600359 11/13/2016 17:59:58

0.100 18:02

0.099 18:07

Test Message: 0.13

Lab #: 072600360 11/16/2016 02:56:02

0.100 02:58

0.098 03:02

Test Message: 0.19

Lab #: 072600361 11/19/2016 21:20:08

0.099 21:24

0.099 21:30

Test Message: 0.04

Lab #: 072600362 11/20/2016 00:00:25

0.099 00:03

0.098 00:09

Test Message: 0.15

Lab #: 072600363 11/20/2016 01:47:51

0.098 01:50

0.098 01:56

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000726

GEBS Results

GEBS VALUE = 0.100

Test Message: 0.15

INTOXILYZER 8000
Instrument Initialization
21:36 11/25/16

SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000726
11/25/16 21:42

Lab ID# = 072600364
GEBS Value = 0.100 g/210L
The U of M is +/- 0.006 (95% CI)

Charge = 23152A
15 min observation before sequence? Yes

Test	g/210L	Time
Air Blank	0.000	21:43
Diagnostic Test	Pass	21:44
Air Blank	0.000	21:44
GEBS Check	0.101	21:44
Air Blank	0.000	21:45
Subject Test	0.167	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Test	0.181	21:49
Air Blank	0.000	21:49
GEBS Check	0.099	21:50
Air Blank	0.000	21:50
Diagnostic Test	Pass	21:50
Air Blank	0.000	21:51

Successfully
Completed Test

Reported Result: .16/.18

Improbable Admission Given

INTOXILYZER 8000
Instrument Initialization
23:28 11/25/16

SAN DIEGO PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000726
11/25/16 23:47

Lab ID# = 072603365
GEBS Value = 0.100 g/210L
The U of M is +/-0.006 (95% CI)

Charge = 23152A
15 min observation before sequence? Yes

Test	g/210L	Time
Air Blank	0.000	23:48
Diagnostic Test	Pass	23:48
Air Blank	0.000	23:49
GEBS Check	0.101	23:49
Air Blank	0.003	23:49
Subject Test	0.113	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:52
Subject Test	0.112	23:53
Air Blank	0.000	23:54
GEBS Check	0.101	23:54
Air Blank	0.000	23:55
Diagnostic Test	Pass	23:55
Air Blank	0.000	23:55

Successfully
Completed Test

Reported Result: .11/.11

Troncetta/Admonishment Given

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

Lab #: 072703607 11/01/2016 02:07:19

0.097 02:09

0.097 02:15

Test Message: 0.12

Lab #: 072703608 11/01/2016 03:35:45

0.098 03:37

0.095 03:44

Test Message: 0.19

Lab #: 072703609 11/09/2016 21:07:32

0.098 21:10

0.097 21:15

Test Message: 0.08

Lab #: 072703610 11/10/2016 04:03:37

0.098 04:06

0.097 04:11

Test Message: 0.15

Lab #: 072703611 11/10/2016 22:27:08

0.098 22:29

Test Message: Subject Test Refused

Lab #: 072703612 11/14/2016 19:38:39

0.098 19:41

0.095 19:47

Test Message: 0.22

Lab #: 072703613 11/14/2016 22:51:54

0.097 22:54

Test Message: Sequence Aborted

Lab #: 072703614 11/19/2016 00:15:38

0.098 00:18

0.095 00:26

Test Message: 0.16

Lab #: 072703615 11/20/2016 22:46:25

0.098 23:05

0.098 23:10

Test Message: 0.19

Lab #: 072703616 11/25/2016 03:55:20

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

Lab #: 072703616 11/25/2016 03:55:20

0.099 03:59

0.096 04:06

Test Message: 0.12

Lab #: 072703617 11/30/2016 11:52:05

0.099 11:54

0.095 12:01

Test Message: 0.31

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000728

GEBS Results

GEBS VALUE = 0.100

Lab #: 072802441 11/04/2016 23:59:12

0.099 00:01

Test Message: Deficient Sample

Lab #: 072802442 11/06/2016 22:46:36

0.100 22:48

0.098 22:56

Test Message: 0.14

Lab #: 072802443 11/13/2016 01:25:30

0.100 01:27

0.096 01:35

Test Message: 0.22

Lab #: 072802444 11/13/2016 23:17:00

0.099 23:19

0.099 23:24

Test Message: 0.14

Lab #: 072802445 11/20/2016 01:49:21

0.099 01:51

0.098 01:57

Test Message: 0.16

Lab #: 072802446 11/23/2016 23:00:44

0.100 23:02

0.097 23:10

Test Message: 0.24

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500739 11/13/2016 01:05:01

0.099 01:07

0.098 01:12

Test Message: 0.14

Lab #: 605500740 11/13/2016 01:28:13

0.098 01:30

0.097 01:37

Test Message: 0.18

Lab #: 605500741 11/18/2016 23:47:54

0.100 23:50

0.097 23:57

Test Message: 0.18

Lab #: 605500742 11/19/2016 01:21:00

0.099 01:23

0.098 01:27

Test Message: 0.17

Lab #: 605500743 11/19/2016 02:28:45

0.099 02:31

0.098 02:36

Test Message: 0.23

Lab #: 605500744 11/19/2016 02:43:15

0.099 02:45

0.097 02:50

Test Message: 0.16

Lab #: 605500745 11/19/2016 22:35:44

0.098 22:39

0.097 22:44

Test Message: 0.21

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801339 11/01/2016 01:52:29

0.099 01:54

0.095 01:59

Test Message: 0.15

Lab #: 605801340 11/01/2016 03:16:26

0.098 03:18

Test Message: Radio Frequency Interferent Detected

Lab #: 605801341 11/03/2016 14:59:02

0.099 15:01

Test Message: Deficient Sample

Lab #: 605801342 11/03/2016 15:06:38

0.096 15:08

Test Message: No Sample Given

Lab #: 605801343 11/05/2016 03:38:58

0.098 03:41

0.097 03:48

Test Message: 0.10

Lab #: 605801344 11/05/2016 21:45:55

0.098 21:50

0.097 21:58

Test Message: 0.18

Lab #: 605801345 11/05/2016 22:06:45

0.098 22:09

0.095 22:15

Test Message: 0.25

Lab #: 605801346 11/06/2016 00:54:13

0.098 00:56

0.097 01:01

Test Message: 0.24

Lab #: 605801347 11/06/2016 16:24:43

0.098 16:27

Test Message: Deficient Sample

Lab #: 605801348 11/06/2016 16:38:03

0.097 16:40

0.096 16:46

Test Message: 0.15

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801349 11/06/2016 20:09:27

0.098 20:11

0.096 20:17

Test Message: 0.21

Lab #: 605801350 11/07/2016 23:13:25

0.098 23:15

0.097 23:21

Test Message: 0.10

Lab #: 605801351 11/09/2016 22:51:59

0.099 22:54

Test Message: Deficient Sample

Lab #: 605801352 11/10/2016 15:25:24

0.099 15:28

Test Message: Radio Frequency Interferent Detected

Lab #: 605801353 11/10/2016 15:33:51

0.099 15:36

Test Message: Wait Fifteen Minutes; Repeat Test

Lab #: 605801354 11/10/2016 15:38:29

0.098 15:40

0.097 15:45

Test Message: 0.32

Lab #: 605801355 11/10/2016 20:09:51

0.098 20:12

Test Message: Subject Test Refused

Lab #: 605801356 11/10/2016 22:25:00

0.099 22:27

Test Message: Interferent Detected Choose Another Test

Lab #: 605801357 11/11/2016 22:54:48

0.099 22:56

0.097 23:05

Test Message: 0.09

Lab #: 605801358 11/12/2016 00:19:25

0.099 00:25

0.097 00:31

Test Message: 0.09

Lab #: 605801359 11/12/2016 01:25:34

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801359 11/12/2016 01:25:34

0.099 01:27

0.097 01:33

Test Message: 0.28

Lab #: 605801360 11/12/2016 03:03:47

0.099 03:05

0.098 03:10

Test Message: 0.19

Lab #: 605801361 11/12/2016 03:30:13

0.099 03:32

0.098 03:38

Test Message: 0.20

Lab #: 605801362 11/12/2016 21:55:56

0.099 21:57

0.098 22:03

Test Message: 0.16

Lab #: 605801363 11/13/2016 04:21:45

0.099 04:24

0.098 04:29

Test Message: 0.18

Lab #: 605801364 11/14/2016 04:29:55

0.099 04:32

0.098 04:37

Test Message: 0.14

Lab #: 605801365 11/14/2016 15:35:36

0.099 15:37

0.098 15:43

Test Message: 0.23

Lab #: 605801366 11/15/2016 00:41:58

0.099 01:00

0.097 01:05

Test Message: 0.20

Lab #: 605801367 11/15/2016 15:00:54

0.099 15:03

0.098 15:09

Test Message: 0.14

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801368 11/15/2016 23:43:57

0.098 23:46

0.097 23:51

Test Message: 0.16

Lab #: 605801369 11/16/2016 03:02:09

0.099 03:04

0.096 03:09

Test Message: 0.16

Lab #: 605801370 11/16/2016 04:57:06

0.099 04:59

0.097 05:04

Test Message: 0.15

Lab #: 605801371 11/16/2016 20:17:30

0.098 20:20

0.097 20:26

Test Message: 0.18

Lab #: 605801372 11/17/2016 04:44:18

0.099 04:46

0.099 04:51

Test Message: 0.07

Lab #: 605801373 11/17/2016 22:19:19

0.099 22:21

0.098 22:26

Test Message: 0.23

Lab #: 605801374 11/18/2016 01:12:02

0.101 01:14

0.100 01:19

Test Message: 0.19

Lab #: 605801375 11/19/2016 20:29:59

0.098 20:32

0.096 20:37

Test Message: 0.23

Lab #: 605801376 11/20/2016 11:17:50

0.098 11:21

0.097 11:26

Test Message: 0.19

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801377 11/20/2016 20:45:24

0.099 20:47

Test Message: Subject Test Refused

Lab #: 605801378 11/20/2016 22:26:07

0.099 22:28

0.098 22:33

Test Message: 0.14

Lab #: 605801379 11/21/2016 19:46:14

0.099 19:52

0.096 19:59

Test Message: 0.19

Lab #: 605801380 11/22/2016 19:29:59

0.099 19:32

0.097 19:37

Test Message: 0.36

Lab #: 605801381 11/23/2016 03:56:25

0.099 03:58

0.097 04:04

Test Message: 0.16

Lab #: 605801382 11/23/2016 17:15:40

0.099 17:18

0.096 17:23

Test Message: 0.17

Lab #: 605801383 11/23/2016 19:35:36

0.099 19:37

0.099 19:43

Test Message: 0.05

Lab #: 605801384 11/23/2016 23:47:34

0.099 23:50

0.098 23:56

Test Message: 0.10

Lab #: 605801385 11/25/2016 02:11:24

0.100 02:14

0.098 02:20

Test Message: 0.09

Lab #: 605801386 11/25/2016 03:50:04

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801386 11/25/2016 03:50:04

0.100 03:52

Test Message: Radio Frequency Interferent Detected

Lab #: 605801387 11/25/2016 03:54:50

0.099 03:56

0.098 04:01

Test Message: 0.19

Lab #: 605801388 11/25/2016 20:08:04

0.100 20:10

0.098 20:15

Test Message: 0.17

Lab #: 605801389 11/26/2016 00:38:44

0.098 00:42

0.099 00:47

Test Message: 0.14

Lab #: 605801390 11/26/2016 04:42:59

0.101 05:02

0.099 05:07

Test Message: 0.09

Lab #: 605801391 11/27/2016 12:57:45

0.099 13:00

0.098 13:05

Test Message: 0.14

Lab #: 605801392 11/28/2016 02:35:09

0.100 02:37

0.097 02:44

Test Message: 0.19

Lab #: 605801393 11/28/2016 18:24:31

0.099 18:27

0.097 18:32

Test Message: 0.17

Lab #: 605801394 11/28/2016 23:34:11

0.100 23:36

0.097 23:41

Test Message: 0.19

Lab #: 605801395 11/29/2016 01:20:41

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801395 11/29/2016 01:20:41

0.099 01:22

Test Message: Radio Frequency Interferent Detected

Lab #: 605801396 11/30/2016 05:09:57

0.100 05:12

0.098 05:17

Test Message: 0.14

Lab #: 605801397 11/30/2016 11:46:34

0.100 11:48

Test Message: Radio Frequency Interferent Detected

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902696 11/13/2016 02:16:45

0.098 02:19

0.097 02:25

Test Message: 0.07

Lab #: 605902697 11/18/2016 23:59:05

0.099 00:01

0.098 00:06

Test Message: 0.10

Lab #: 605902698 11/19/2016 01:20:28

0.098 01:22

0.098 01:27

Test Message: 0.11

Lab #: 605902699 11/19/2016 02:25:00

0.099 02:27

0.098 02:34

Test Message: 0.08

Lab #: 605902700 11/19/2016 02:36:57

0.099 02:38

0.099 02:44

Test Message: 0.10

Lab #: 605902701 11/19/2016 02:58:33

0.099 03:01

0.098 03:07

Test Message: 0.15

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000726

GEBS Results

GEBS VALUE = 0.100

Lab #: 072600352 10/04/2016 01:55:41

0.100 01:57

Test Message: Sequence Aborted

Lab #: 072600353 10/19/2016 03:09:37

0.099 03:11

0.098 03:16

Test Message: 0.18

Lab #: 072600354 10/31/2016 23:18:23

0.099 23:21

0.099 23:27

Test Message: 0.27

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000719

GEBS Results

GEBS VALUE = 0.100

Lab #: 071904212 10/03/2016 23:46:41

0.099 23:48

0.098 23:53

Test Message: 0.13

Lab #: 071904213 10/04/2016 21:36:30

0.099 21:39

0.097 21:44

Test Message: 0.19

Lab #: 071904214 10/05/2016 00:46:47

0.099 00:50

0.098 00:55

Test Message: 0.11

Lab #: 071904215 10/12/2016 03:04:29

0.099 03:07

0.098 03:12

Test Message: 0.14

Lab #: 071904216 10/13/2016 21:32:48

0.100 21:35

0.098 21:39

Test Message: 0.20

Lab #: 071904217 10/15/2016 01:11:37

0.099 01:13

Test Message: Deficient Sample

Lab #: 071904219 10/20/2016 01:36:11

0.099 01:38

0.098 01:43

Test Message: 0.17

Lab #: 071904220 10/22/2016 00:44:29

0.100 00:46

Test Message: Sequence Aborted

Lab #: 071904221 10/28/2016 03:06:48

0.098 03:08

0.098 03:13

Test Message: 0.11

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

Lab #: 072703588 10/01/2016 01:54:12

0.097 01:57

0.094 02:07

Test Message: 0.12

Lab #: 072703589 10/02/2016 11:04:40

0.097 11:12

0.096 11:17

Test Message: 0.16

Lab #: 072703590 10/09/2016 18:34:19

0.098 18:36

0.096 18:42

Test Message: 0.13

Lab #: 072703591 10/13/2016 18:07:17

0.099 18:10

0.096 18:14

Test Message: 0.10

Lab #: 072703593 10/15/2016 01:47:08

0.097 01:49

0.096 01:54

Test Message: 0.11

Lab #: 072703594 10/16/2016 09:26:40

0.097 09:29

0.095 09:34

Test Message: 0.17

Lab #: 072703595 10/19/2016 04:14:33

0.098 04:17

0.096 04:22

Test Message: 0.10

Lab #: 072703596 10/23/2016 08:16:15

0.098 08:19

0.095 08:24

Test Message: 0.03

Lab #: 072703598 10/23/2016 23:03:42

0.097 23:06

Test Message: Radio Frequency Interferent Detected

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

Lab #: 072703599 10/23/2016 23:07:30

0.098 23:09

0.095 23:14

Test Message: 0.12

Lab #: 072703600 10/23/2016 23:20:21

0.096 23:24

0.095 23:29

Test Message: 0.21

Lab #: 072703601 10/23/2016 23:33:17

0.097 23:35

0.095 23:42

Test Message: 0.11

Lab #: 072703602 10/25/2016 21:32:16

0.098 21:34

0.095 21:40

Test Message: 0.12

Lab #: 072703603 10/28/2016 02:44:24

0.098 02:48

Test Message: No Sample Given

Lab #: 072703604 10/29/2016 04:13:23

0.098 04:15

0.096 04:20

Test Message: 0.18

Lab #: 072703605 10/30/2016 03:40:55

0.098 03:43

Test Message: Deficient Sample

Lab #: 072703606 10/30/2016 03:53:16

0.096 03:55

0.095 04:00

Test Message: 0.18

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000728

GEBS Results

GEBS VALUE = 0.100

Lab #: 072802432 10/04/2016 03:13:44

0.100 03:15

0.099 03:20

Test Message: 0.21

Lab #: 072802433 10/09/2016 21:04:30

0.100 21:07

0.098 21:11

Test Message: Diagnostic Failed

Lab #: 072802434 10/12/2016 02:25:52

0.100 02:27

0.098 02:33

Test Message: 0.05

Lab #: 072802435 10/13/2016 21:39:00

0.099 21:41

Test Message: Sequence Aborted

Lab #: 072802437 10/13/2016 22:39:04

0.099 22:41

0.098 22:45

Test Message: 0.21

Lab #: 072802438 10/17/2016 00:44:31

0.100 00:46

Test Message: Deficient Sample

Lab #: 072802439 10/19/2016 02:18:03

0.099 02:20

0.099 02:25

Test Message: 0.08

Lab #: 072802440 10/30/2016 03:01:05

0.099 03:03

0.099 03:08

Test Message: 0.10

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500730 10/22/2016 00:26:13

0.099 00:28

0.098 00:33

Test Message: 0.28

Lab #: 605500731 10/22/2016 01:19:45

0.098 01:21

0.099 01:26

Test Message: 0.09

Lab #: 605500732 10/22/2016 01:39:30

0.098 01:41

0.098 01:46

Test Message: 0.16

Lab #: 605500733 10/22/2016 01:51:23

0.098 01:54

0.098 01:58

Test Message: 0.09

Lab #: 605500734 10/22/2016 02:24:37

0.098 02:27

0.096 02:31

Test Message: 0.12

Lab #: 605500735 10/30/2016 00:51:21

0.099 00:54

0.098 00:59

Test Message: 0.08

Lab #: 605500736 10/30/2016 01:03:43

0.098 01:06

Test Message: Sequence Aborted

Lab #: 605500737 10/30/2016 01:53:24

0.098 01:55

0.097 02:01

Test Message: 0.12

Lab #: 605500738 10/30/2016 02:43:45

0.098 02:45

0.097 02:50

Test Message: 0.13

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801265 10/01/2016 04:21:19

0.099 04:24

0.098 04:30

Test Message: 0.13

Lab #: 605801266 10/01/2016 15:50:59

0.098 15:53

0.096 16:01

Test Message: 0.12

Lab #: 605801267 10/01/2016 16:22:55

0.099 16:25

Test Message: Subject Test Refused

Lab #: 605801268 10/01/2016 21:43:07

0.099 21:45

0.098 21:50

Test Message: 0.20

Lab #: 605801269 10/03/2016 19:42:42

0.099 19:45

0.098 19:50

Test Message: 0.27

Lab #: 605801270 10/03/2016 21:11:04

0.100 21:13

0.099 21:18

Test Message: 0.06

Lab #: 605801271 10/04/2016 03:29:16

0.099 03:31

Test Message: Subject Test Refused

Lab #: 605801272 10/04/2016 03:35:02

0.098 03:37

0.100 03:42

Test Message: 0.19

Lab #: 605801273 10/05/2016 05:45:22

0.099 05:47

0.098 05:52

Test Message: 0.18

Lab #: 605801274 10/05/2016 21:18:12

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801274 10/05/2016 21:18:12

0.099 21:20

0.098 21:25

Test Message: 0.06

Lab #: 605801275 10/06/2016 00:00:46

0.099 00:02

0.098 00:07

Test Message: 0.12

Lab #: 605801276 10/06/2016 03:10:05

0.099 03:15

0.097 03:21

Test Message: 0.23

Lab #: 605801277 10/06/2016 03:29:41

0.099 03:32

0.098 03:39

Test Message: 0.17

Lab #: 605801278 10/07/2016 02:33:24

0.099 02:35

0.099 02:41

Test Message: 0.11

Lab #: 605801279 10/07/2016 02:50:58

0.099 02:54

Test Message: Subject Test Refused

Lab #: 605801280 10/08/2016 01:17:24

0.098 01:20

Test Message: No Sample Given

Lab #: 605801281 10/08/2016 01:33:23

0.098 01:36

Test Message: Deficient Sample

Lab #: 605801282 10/08/2016 02:28:33

0.099 02:30

0.098 02:36

Test Message: 0.12

Lab #: 605801283 10/08/2016 04:23:54

0.100 04:26

Test Message: Deficient Sample

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801284 10/08/2016 04:31:54

0.099 04:33

Test Message: Deficient Sample

Lab #: 605801285 10/08/2016 04:39:53

0.098 04:41

Test Message: Deficient Sample

Lab #: 605801286 10/08/2016 04:47:15

0.098 04:49

Test Message: Deficient Sample

Lab #: 605801287 10/08/2016 04:58:26

0.098 05:00

Test Message: Radio Frequency Interferent Detected

Lab #: 605801288 10/08/2016 05:03:23

0.099 05:05

0.097 05:11

Test Message: 0.14

Lab #: 605801289 10/08/2016 15:54:38

0.098 15:56

Test Message: Sequence Aborted

Lab #: 605801290 10/08/2016 23:48:46

0.099 23:52

0.098 23:59

Test Message: 0.07

Lab #: 605801291 10/09/2016 16:18:50

0.099 16:21

0.099 16:26

Test Message: 0.18

Lab #: 605801292 10/09/2016 23:34:04

0.099 23:36

0.097 23:43

Test Message: 0.18

Lab #: 605801293 10/10/2016 22:36:32

0.098 22:40

0.097 22:45

Test Message: 0.12

Lab #: 605801294 10/12/2016 02:54:07

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801294 10/12/2016 02:54:07

0.099 02:56

Test Message: Subject Test Refused

Lab #: 605801295 10/12/2016 16:48:55

0.098 16:51

0.096 16:56

Test Message: 0.25

Lab #: 605801296 10/13/2016 16:19:04

0.099 16:21

0.096 16:28

Test Message: 0.18

Lab #: 605801297 10/13/2016 20:36:24

0.099 20:54

0.096 21:00

Test Message: 0.13

Lab #: 605801298 10/13/2016 23:34:00

0.098 23:36

0.097 23:41

Test Message: 0.15

Lab #: 605801299 10/14/2016 03:33:46

0.098 03:36

0.097 03:42

Test Message: 0.16

Lab #: 605801300 10/14/2016 23:30:36

0.098 23:32

0.097 23:39

Test Message: 0.10

Lab #: 605801301 10/15/2016 00:21:03

0.098 00:24

0.098 00:29

Test Message: 0.12

Lab #: 605801302 10/15/2016 02:32:24

0.099 02:34

Test Message: Subject Test Refused

Lab #: 605801303 10/15/2016 04:16:40

0.100 04:19

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801303 10/15/2016 04:16:40

0.098 04:25

Test Message: 0.16

Lab #: 605801304 10/15/2016 19:52:55

0.099 19:55

0.098 20:00

Test Message: 0.05

Lab #: 605801305 10/15/2016 21:34:24

0.100 21:37

Test Message: Deficient Sample

Lab #: 605801306 10/15/2016 21:41:41

0.097 21:44

Test Message: Deficient Sample

Lab #: 605801307 10/15/2016 21:48:43

0.098 21:50

Test Message: Deficient Sample

Lab #: 605801308 10/15/2016 22:01:32

0.097 22:03

Test Message: Deficient Sample

Lab #: 605801309 10/16/2016 17:42:38

0.099 17:44

0.098 17:49

Test Message: 0.26

Lab #: 605801310 10/16/2016 17:55:36

0.098 17:57

Test Message: Deficient Sample

Lab #: 605801311 10/16/2016 18:02:29

0.098 18:04

0.096 18:11

Test Message: 0.10

Lab #: 605801312 10/16/2016 20:02:25

0.100 20:04

Test Message: Deficient Sample

Lab #: 605801313 10/17/2016 01:55:12

0.098 01:57

0.098 02:02

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Test Message: 0.13

Lab #: 605801314 10/17/2016 21:18:11

0.099 21:20

0.096 21:28

Test Message: 0.27

Lab #: 605801315 10/18/2016 23:06:07

0.100 23:08

0.096 23:14

Test Message: 0.23

Lab #: 605801316 10/20/2016 06:32:57

0.100 06:36

0.099 06:41

Test Message: 0.19

Lab #: 605801317 10/21/2016 01:46:30

0.101 01:49

0.100 01:54

Test Message: 0.13

Lab #: 605801318 10/21/2016 02:39:56

0.100 02:42

0.099 02:48

Test Message: 0.20

Lab #: 605801319 10/21/2016 22:38:46

0.100 22:40

0.098 22:45

Test Message: 0.18

Lab #: 605801320 10/22/2016 01:22:54

0.099 01:24

0.098 01:30

Test Message: 0.09

Lab #: 605801321 10/22/2016 05:00:35

0.099 05:03

0.099 05:08

Test Message: 0.09

Lab #: 605801322 10/23/2016 01:56:09

0.099 01:58

0.099 02:03

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Test Message: 0.10

Lab #: 605801323 10/23/2016 05:19:32

0.099 05:21

0.098 05:26

Test Message: 0.18

Lab #: 605801324 10/23/2016 18:41:46

0.099 18:43

0.098 18:48

Test Message: 0.16

Lab #: 605801325 10/24/2016 02:17:46

0.097 02:20

Test Message: No Sample Given

Lab #: 605801326 10/24/2016 02:40:33

0.098 02:43

Test Message: No Sample Given

Lab #: 605801327 10/24/2016 20:58:34

0.097 21:00

0.097 21:05

Test Message: 0.21

Lab #: 605801328 10/25/2016 02:30:46

0.098 02:33

Test Message: Subject Test Refused

Lab #: 605801329 10/27/2016 05:11:19

0.098 05:13

0.097 05:18

Test Message: 0.22

Lab #: 605801330 10/28/2016 23:38:34

0.098 23:41

0.097 23:46

Test Message: 0.18

Lab #: 605801331 10/29/2016 04:16:08

0.098 04:18

0.097 04:23

Test Message: 0.24

Lab #: 605801332 10/29/2016 23:21:25

0.098 23:23

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801332 10/29/2016 23:21:25

0.095 23:30

Test Message: 0.21

Lab #: 605801333 10/30/2016 01:42:21

0.098 01:44

0.095 01:50

Test Message: 0.15

Lab #: 605801334 10/30/2016 02:14:16

0.098 02:17

0.097 02:22

Test Message: 0.18

Lab #: 605801335 10/30/2016 03:42:26

0.098 03:44

0.096 03:52

Test Message: 0.08

Lab #: 605801336 10/31/2016 01:14:16

0.099 01:16

Test Message: Radio Frequency Interferent Detected

Lab #: 605801337 10/31/2016 01:17:53

0.098 01:20

0.096 01:25

Test Message: 0.18

Lab #: 605801338 10/31/2016 03:53:26

0.098 03:56

0.095 04:02

Test Message: 0.23

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902683 10/08/2016 01:53:37

0.097 01:56

0.096 02:01

Test Message: 0.23

Lab #: 605902684 10/22/2016 00:34:53

0.099 00:37

0.098 00:42

Test Message: 0.09

Lab #: 605902685 10/22/2016 01:07:31

0.098 01:10

0.098 01:16

Test Message: 0.09

Lab #: 605902686 10/22/2016 01:40:47

0.097 01:43

0.097 01:48

Test Message: 0.17

Lab #: 605902687 10/22/2016 02:26:50

0.097 02:29

0.097 02:34

Test Message: 0.12

Lab #: 605902688 10/22/2016 02:57:25

0.099 02:59

0.098 03:03

Test Message: 0.10

Lab #: 605902689 10/29/2016 23:59:33

0.097 00:01

0.098 00:07

Test Message: 0.09

Lab #: 605902690 10/30/2016 00:52:53

0.098 00:55

Test Message: Sequence Aborted

Lab #: 605902691 10/30/2016 01:01:08

0.098 01:03

0.097 01:08

Test Message: 0.11

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902692 10/30/2016 01:52:22

0.097 01:54

0.095 02:00

Test Message: 0.21

Lab #: 605902693 10/30/2016 02:23:53

0.097 02:25

Test Message: Deficient Sample

Lab #: 605902694 10/30/2016 02:30:58

0.097 02:32

0.095 02:38

Test Message: 0.14

Lab #: 605902695 10/30/2016 03:09:28

0.098 03:11

0.096 03:17

Test Message: 0.22

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000718

GEBS Results

GEBS VALUE = 0.100

Lab #: 071801726 09/06/2016 01:34:08

0.102 01:36

0.099 01:41

Test Message: 0.11

Lab #: 071801727 09/06/2016 23:49:00

0.100 23:51

0.098 23:57

Test Message: 0.12

Lab #: 071801728 09/09/2016 18:51:42

0.099 18:54

0.099 18:59

Test Message: 0.12

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000719

GEBS Results

GEBS VALUE = 0.100

Lab #: 071904205 09/19/2016 14:11:42

0.099 14:13

0.098 14:17

Test Message: 0.00

Lab #: 071904206 09/20/2016 00:42:19

0.099 00:44

0.098 00:49

Test Message: 0.12

Lab #: 071904207 09/20/2016 22:49:06

0.099 22:51

0.096 22:57

Test Message: 0.15

Lab #: 071904208 09/22/2016 00:50:30

0.099 00:52

0.098 00:56

Test Message: 0.14

Lab #: 071904209 09/23/2016 18:21:13

0.099 18:40

Test Message: Sequence Aborted

Lab #: 071904210 09/27/2016 00:12:25

0.099 00:15

0.098 00:20

Test Message: 0.16

Lab #: 071904211 09/29/2016 01:56:43

0.099 01:58

0.098 02:03

Test Message: 0.10

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000726

GEBS Results

GEBS VALUE = 0.100

Lab #: 072600349 09/18/2016 20:40:54

0.099 20:43

0.098 20:48

Test Message: 0.07

Lab #: 072600350 09/24/2016 00:31:14

0.099 00:33

0.099 00:38

Test Message: 0.08

Lab #: 072600351 09/25/2016 00:12:37

0.099 00:15

0.096 00:20

Test Message: 0.23

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

Lab #: 072703561 09/01/2016 06:22:55

0.098 06:25

Test Message: No Sample Given

Lab #: 072703562 09/01/2016 07:12:29

0.098 07:14

Test Message: Deficient Sample

Lab #: 072703563 09/01/2016 07:20:16

0.098 07:22

Test Message: No Sample Given

Lab #: 072703564 09/02/2016 00:59:15

0.099 01:01

0.097 01:07

Test Message: 0.16

Lab #: 072703565 09/04/2016 00:18:05

0.098 00:20

0.097 00:25

Test Message: 0.22

Lab #: 072703566 09/04/2016 02:28:32

0.098 02:31

Test Message: No Sample Given

Lab #: 072703567 09/04/2016 02:36:24

0.098 02:38

Test Message: Sequence Aborted

Lab #: 072703568 09/08/2016 01:30:31

0.098 01:32

0.097 01:38

Test Message: 0.12

Lab #: 072703569 09/08/2016 01:43:18

0.098 01:46

0.097 01:51

Test Message: 0.15

Lab #: 072703570 09/09/2016 02:28:22

0.097 02:30

Test Message: Radio Frequency Interferent Detected

Lab #: 072703571 09/09/2016 02:32:29

0.098 02:34

SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

Lab #: 072703571 09/09/2016 02:32:29

0.096 02:39

Test Message: 0.14

Lab #: 072703572 09/10/2016 01:29:55

0.096 01:32

0.094 01:37

Test Message: 0.15

Lab #: 072703573 09/11/2016 03:16:40

0.098 03:18

Test Message: Radio Frequency Interferent Detected

Lab #: 072703574 09/11/2016 03:21:18

0.098 03:23

0.096 03:30

Test Message: 0.15

Lab #: 072703575 09/12/2016 01:47:34

0.099 01:51

0.097 01:57

Test Message: 0.15

Lab #: 072703576 09/12/2016 23:56:59

0.099 23:58

0.098 00:05

Test Message: 0.08

Lab #: 072703577 09/15/2016 04:32:53

0.099 04:36

0.095 04:44

Test Message: 0.17

Lab #: 072703578 09/17/2016 03:22:10

0.098 03:24

0.097 03:30

Test Message: 0.14

Lab #: 072703579 09/18/2016 00:38:51

0.097 00:41

0.097 00:47

Test Message: 0.18

Lab #: 072703580 09/18/2016 18:37:25

0.097 18:39