

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000718

GEBS Results

GEBS VALUE = 0.100

Lab #: 071801693 06/05/2016 23:43:09

0.100 23:45

0.097 23:50

Test Message: 0.18

Lab #: 071801694 06/08/2016 01:46:10

0.102 01:48

0.100 01:52

Test Message: 0.19

Lab #: 071801695 06/11/2016 02:19:52

0.101 02:23

0.099 02:28

Test Message: 0.08

Lab #: 071801696 06/12/2016 00:20:29

0.101 00:22

Test Message: Sequence Aborted

Lab #: 071801697 06/12/2016 23:39:36

0.101 23:41

Test Message: Sequence Aborted

Lab #: 071801698 06/24/2016 21:51:37

0.100 21:53

0.101 21:58

Test Message: 0.16

Lab #: 071801699 06/24/2016 23:56:53

0.100 23:59

Test Message: Sequence Aborted

Lab #: 071801700 06/25/2016 00:04:57

0.100 00:06

0.098 00:11

Test Message: 0.21

Lab #: 071801701 06/27/2016 00:31:12

0.000 00:34

Test Message: Improper Sample

Lab #: 071801702 06/27/2016 00:35:05

0.102 00:37

Test Message: Deficient Sample

Lab #: 071801703 06/27/2016 00:41:58

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000718

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 071801703 06/27/2016 00:41:58

0.100 00:43

0.099 00:50

Test Message: 0.15

Lab #: 071801704 06/27/2016 20:40:34

0.100 20:42

0.097 20:48

Test Message: 0.29

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000726

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 072600329 06/04/2016 02:50:49

0.099 02:52

Test Message: No Sample Given

Lab #: 072600330 06/04/2016 02:57:13

0.098 02:59

0.098 03:06

Test Message: 0.12

Lab #: 072600331 06/05/2016 01:14:28

0.099 01:17

0.097 01:25

Test Message: 0.18

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000728

GEBS Results

GEBS VALUE = 0.100

Lab #: 072802403 06/13/2016 22:42:30

0.101 22:44

0.099 22:52

Test Message: 0.11

Lab #: 072802404 06/14/2016 00:21:48

0.101 00:23

0.099 00:30

Test Message: 0.11

Lab #: 072802405 06/18/2016 23:44:05

0.100 23:46

0.100 23:52

Test Message: 0.13

Lab #: 072802406 06/19/2016 02:00:50

0.101 02:03

0.100 02:09

Test Message: 0.20

Lab #: 072802407 06/20/2016 02:30:30

0.101 02:33

0.100 02:39

Test Message: 0.13

Lab #: 072802408 06/24/2016 21:20:17

0.101 21:22

0.098 21:29

Test Message: 0.21

Lab #: 072802409 06/25/2016 01:48:32

0.100 01:50

0.100 01:55

Test Message: 0.15

Lab #: 072802410 06/26/2016 01:28:56

0.100 01:30

Test Message: Sequence Aborted

Lab #: 072802411 06/26/2016 21:44:42

0.100 21:46

Test Message: Sequence Aborted

Lab #: 072802412 06/26/2016 22:05:36

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000728

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 072802412 06/26/2016 22:05:36

0.100 22:07

0.100 22:12

Test Message: 0.06

Lab #: 072802414 06/27/2016 02:00:25

0.100 02:02

0.099 02:08

Test Message: 0.10

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500654 06/03/2016 23:56:53

0.100 23:58

0.099 00:04

Test Message: 0.00

Lab #: 605500655 06/10/2016 23:18:47

0.100 23:20

0.100 23:26

Test Message: 0.02

Lab #: 605500656 06/11/2016 00:09:02

0.100 00:11

0.099 00:16

Test Message: 0.10

Lab #: 605500657 06/11/2016 00:22:59

0.100 00:25

0.099 00:30

Test Message: 0.11

Lab #: 605500658 06/11/2016 01:23:31

0.100 01:25

0.098 01:31

Test Message: 0.12

Lab #: 605500659 06/11/2016 02:13:37

0.099 02:15

0.098 02:21

Test Message: 0.10

Lab #: 605500660 06/11/2016 02:31:11

0.098 02:33

0.098 02:39

Test Message: 0.11

Lab #: 605500661 06/11/2016 03:13:20

0.099 03:15

0.098 03:20

Test Message: 0.14

Lab #: 605500662 06/18/2016 00:18:42

0.100 00:20

0.100 00:26

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Test Message: 0.03

Lab #: 605500663 06/18/2016 01:00:29

0.100 01:02

0.099 01:07

Test Message: 0.09

Lab #: 605500664 06/18/2016 01:44:50

0.100 01:47

0.099 01:53

Test Message: 0.10

Lab #: 605500665 06/18/2016 02:48:28

0.099 02:50

0.098 03:00

Test Message: 0.11

Lab #: 605500666 06/18/2016 03:21:24

0.098 03:23

Test Message: Sequence Aborted

Lab #: 605500667 06/18/2016 03:37:36

0.099 03:39

0.097 03:46

Test Message: 0.09

Lab #: 605500668 06/23/2016 23:32:53

0.099 23:35

0.097 23:45

Test Message: 0.18

Lab #: 605500669 06/24/2016 00:24:23

0.100 00:26

Test Message: Sequence Aborted

Lab #: 605500670 06/24/2016 00:57:55

0.100 00:59

0.098 01:04

Test Message: 0.14

Lab #: 605500671 06/24/2016 01:34:10

0.099 01:36

0.099 01:42

Test Message: 0.19

Lab #: 605500673 06/24/2016 03:06:56

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500673 06/24/2016 03:06:56

0.099 03:08

Test Message: Deficient Sample

Lab #: 605500674 06/24/2016 03:13:19

0.099 03:15

Test Message: Interferent Detected Choose Another Test

Lab #: 605500675 06/24/2016 03:17:05

0.095 03:18

0.089 03:25

Test Message: CalCheck Failed Return Instrument to Lab

Lab #: 605500676 06/26/2016 01:05:42

0.099 01:07

0.096 01:13

Test Message: 0.14

Lab #: 605500677 06/26/2016 01:36:30

0.099 01:38

0.099 01:43

Test Message: 0.04

Lab #: 605500678 06/26/2016 02:34:32

0.096 02:36

Test Message: Deficient Sample

Lab #: 605500679 06/26/2016 02:43:48

0.098 02:46

0.097 02:52

Test Message: 0.18



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006057

GEBS Results

GEBS VALUE = 0.100

Lab #: 605700507 06/04/2016 00:46:38

0.098 00:49

0.097 00:54

Test Message: 0.11

Lab #: 605700508 06/04/2016 01:24:31

0.097 01:26

0.096 01:30

Test Message: 0.16

Lab #: 605700509 06/04/2016 02:20:37

0.098 02:22

Test Message: Deficient Sample

Lab #: 605700510 06/11/2016 00:13:09

0.098 00:15

0.095 00:20

Test Message: 0.17

Lab #: 605700511 06/11/2016 01:30:41

0.098 01:32

0.096 01:38

Test Message: 0.16

Lab #: 605700512 06/11/2016 02:24:39

0.097 02:27

0.095 02:31

Test Message: 0.11

Lab #: 605700513 06/11/2016 03:21:41

0.098 03:23

0.097 03:29

Test Message: 0.10

Lab #: 605700514 06/18/2016 02:56:13

0.098 02:58

0.097 03:04

Test Message: 0.15

Lab #: 605700515 06/18/2016 03:18:47

0.097 03:21

0.095 03:26

Test Message: 0.20

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006057

GEBS Results

GEBS VALUE = 0.100

Lab #: 605700516 06/24/2016 01:42:26

0.097 01:44

0.097 01:49

Test Message: 0.25

Lab #: 605700517 06/25/2016 23:59:29

0.099 00:01

0.097 00:08

Test Message: 0.11

Lab #: 605700518 06/26/2016 02:25:13

0.097 02:27

0.096 02:31

Test Message: 0.14

Lab #: 605700519 06/26/2016 02:34:18

0.097 02:36

0.096 02:41

Test Message: 0.09

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801064 06/04/2016 01:53:34

0.100 01:55

0.098 02:00

Test Message: 0.22

Lab #: 605801065 06/06/2016 03:25:13

0.099 03:27

0.099 03:31

Test Message: 0.12

Lab #: 605801066 06/07/2016 01:41:00

0.100 01:43

0.097 01:48

Test Message: 0.16

Lab #: 605801068 06/11/2016 23:55:27

0.099 23:57

0.097 00:02

Test Message: 0.23

Lab #: 605801069 06/12/2016 19:02:31

0.100 19:04

0.098 19:10

Test Message: 0.14

Lab #: 605801070 06/19/2016 23:59:22

0.099 00:01

0.097 00:08

Test Message: 0.17

Lab #: 605801071 06/20/2016 03:01:26

0.098 03:03

0.098 03:08

Test Message: 0.20

Lab #: 605801072 06/24/2016 02:22:25

0.100 02:24

0.099 02:30

Test Message: 0.11

Lab #: 605801073 06/27/2016 03:43:52

0.099 03:46

0.096 03:53

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

---

Test Message: 0.24

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902502 06/01/2016 00:39:05

0.099 00:42

0.098 00:47

Test Message: 0.15

Lab #: 605902503 06/01/2016 20:19:40

0.099 20:22

0.097 20:27

Test Message: 0.16

Lab #: 605902504 06/02/2016 00:35:07

0.099 00:40

0.098 00:49

Test Message: 0.13

Lab #: 605902505 06/02/2016 01:05:13

0.099 01:07

0.097 01:12

Test Message: 0.14

Lab #: 605902506 06/02/2016 19:39:51

0.099 19:43

Test Message: Deficient Sample

Lab #: 605902507 06/02/2016 19:47:44

0.098 19:49

0.097 19:57

Test Message: 0.08

Lab #: 605902508 06/02/2016 23:26:15

0.099 23:28

0.097 23:34

Test Message: 0.18

Lab #: 605902509 06/02/2016 23:57:52

0.099 00:00

0.098 00:04

Test Message: 0.15

Lab #: 605902510 06/03/2016 02:28:38

0.099 02:31

0.098 02:36

Test Message: 0.08

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902511 06/03/2016 05:35:44

0.099 05:37

0.096 05:43

Test Message: 0.22

Lab #: 605902512 06/04/2016 01:23:51

0.098 01:26

0.098 01:31

Test Message: 0.11

Lab #: 605902513 06/04/2016 02:02:01

0.099 02:04

0.098 02:09

Test Message: 0.07

Lab #: 605902514 06/05/2016 00:04:22

0.098 00:06

0.097 00:11

Test Message: 0.25

Lab #: 605902515 06/06/2016 20:17:54

0.099 20:20

0.098 20:26

Test Message: 0.16

Lab #: 605902516 06/07/2016 01:45:05

0.099 01:48

0.099 01:54

Test Message: 0.08

Lab #: 605902517 06/07/2016 02:33:35

0.099 02:35

0.099 02:44

Test Message: 0.00

Lab #: 605902518 06/07/2016 03:00:19

0.099 03:03

0.098 03:08

Test Message: 0.08

Lab #: 605902519 06/07/2016 03:20:52

0.099 03:23

0.099 03:29

Test Message: 0.19

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902520 06/09/2016 01:11:05

0.099 01:13

0.105 01:19

Test Message: 0.12

Lab #: 605902521 06/09/2016 01:28:58

0.099 01:31

0.097 01:41

Test Message: 0.18

Lab #: 605902522 06/09/2016 21:50:10

0.099 21:52

0.098 21:57

Test Message: 0.10

Lab #: 605902523 06/10/2016 04:06:43

0.099 04:09

0.097 04:14

Test Message: 0.18

Lab #: 605902524 06/10/2016 23:03:04

0.100 23:05

0.098 23:09

Test Message: 0.18

Lab #: 605902525 06/11/2016 00:26:23

0.099 00:30

0.098 00:35

Test Message: 0.10

Lab #: 605902526 06/11/2016 02:55:59

0.098 02:59

Test Message: Sequence Aborted

Lab #: 605902527 06/11/2016 04:49:12

0.099 04:51

0.097 04:57

Test Message: 0.16

Lab #: 605902528 06/11/2016 16:36:49

0.099 16:38

0.098 16:44

Test Message: 0.11

Lab #: 605902529 06/11/2016 19:21:51

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902529 06/11/2016 19:21:51

0.099 19:23

Test Message: No Sample Given

Lab #: 605902530 06/11/2016 20:23:08

0.098 20:25

0.099 20:30

Test Message: 0.06

Lab #: 605902531 06/12/2016 00:45:31

0.099 00:47

0.099 00:52

Test Message: 0.11

Lab #: 605902532 06/12/2016 03:58:10

0.099 03:59

0.098 04:04

Test Message: 0.20

Lab #: 605902533 06/12/2016 17:01:25

0.099 17:03

Test Message: Deficient Sample

Lab #: 605902534 06/12/2016 17:12:38

0.098 17:14

Test Message: Deficient Sample

Lab #: 605902535 06/12/2016 17:19:12

0.098 17:21

Test Message: Deficient Sample

Lab #: 605902536 06/12/2016 23:12:38

0.098 23:14

0.096 23:19

Test Message: 0.21

Lab #: 605902537 06/13/2016 00:37:21

0.098 00:40

0.097 00:45

Test Message: 0.28

Lab #: 605902538 06/13/2016 16:32:28

0.099 16:34

0.097 16:40

Test Message: 0.15



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902539 06/13/2016 19:12:40

0.099 19:14

0.097 19:19

Test Message: 0.13

Lab #: 605902540 06/15/2016 01:08:30

0.099 01:10

0.098 01:16

Test Message: 0.15

Lab #: 605902541 06/16/2016 02:02:16

0.099 02:04

0.098 02:09

Test Message: 0.15

Lab #: 605902542 06/17/2016 00:26:55

0.098 00:30

0.097 00:35

Test Message: 0.15

Lab #: 605902543 06/17/2016 01:09:43

0.099 01:11

0.100 01:17

Test Message: 0.00

Lab #: 605902544 06/17/2016 13:50:39

0.100 13:53

Test Message: Sequence Aborted

Lab #: 605902545 06/17/2016 22:12:44

0.099 22:15

0.098 22:20

Test Message: 0.16

Lab #: 605902546 06/18/2016 02:51:47

0.098 02:54

0.097 03:00

Test Message: 0.19

Lab #: 605902547 06/18/2016 03:23:19

0.099 03:25

0.098 03:31

Test Message: 0.10

Lab #: 605902548 06/18/2016 03:37:15

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902548 06/18/2016 03:37:15

0.099 03:40

0.098 03:45

Test Message: 0.12

Lab #: 605902549 06/18/2016 05:15:36

0.099 05:17

0.097 05:23

Test Message: 0.11

Lab #: 605902550 06/18/2016 23:58:26

0.099 00:00

Test Message: Deficient Sample

Lab #: 605902551 06/19/2016 00:13:07

0.098 00:15

0.098 00:20

Test Message: 0.07

Lab #: 605902552 06/19/2016 02:16:35

0.099 02:19

0.097 02:27

Test Message: 0.18

Lab #: 605902553 06/19/2016 19:37:34

0.098 19:39

Test Message: Deficient Sample

Lab #: 605902554 06/19/2016 19:45:55

0.098 19:47

0.096 19:55

Test Message: 0.25

Lab #: 605902555 06/19/2016 21:15:24

0.098 21:17

0.097 21:22

Test Message: 0.17

Lab #: 605902556 06/19/2016 23:59:25

0.098 00:01

0.096 00:06

Test Message: 0.16

Lab #: 605902557 06/20/2016 01:25:54

0.098 01:28

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902557 06/20/2016 01:25:54

0.096 01:34

Test Message: 0.20

Lab #: 605902558 06/20/2016 04:25:02

0.099 04:27

0.098 04:32

Test Message: 0.07

Lab #: 605902559 06/20/2016 22:23:02

0.099 22:25

Test Message: Sequence Aborted

Lab #: 605902560 06/20/2016 22:39:43

0.099 22:41

0.097 22:47

Test Message: 0.11

Lab #: 605902561 06/21/2016 01:17:45

0.098 01:21

0.098 01:26

Test Message: 0.16

Lab #: 605902562 06/22/2016 03:35:13

0.098 03:37

0.098 03:41

Test Message: 0.12

Lab #: 605902563 06/22/2016 05:08:42

0.099 05:10

Test Message: Deficient Sample

Lab #: 605902564 06/22/2016 05:15:08

0.098 05:17

0.098 05:24

Test Message: 0.00

Lab #: 605902565 06/23/2016 00:13:49

0.099 00:16

0.097 00:21

Test Message: 0.13

Lab #: 605902566 06/24/2016 00:52:20

0.099 00:54

0.097 00:59

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Test Message: 0.11

Lab #: 605902567 06/24/2016 09:24:36

0.099 09:27

0.098 09:32

Test Message: 0.10

Lab #: 605902568 06/25/2016 01:20:05

0.099 01:21

0.096 01:28

Test Message: 0.13

Lab #: 605902569 06/25/2016 03:47:45

0.098 03:49

0.097 03:54

Test Message: 0.07

Lab #: 605902570 06/27/2016 02:31:18

0.098 02:33

Test Message: No Sample Given

Lab #: 605902571 06/27/2016 02:43:37

0.099 02:45

Test Message: Deficient Sample

Lab #: 605902572 06/27/2016 02:52:35

0.098 02:54

0.097 03:00

Test Message: 0.10

Lab #: 605902573 06/27/2016 19:13:31

0.098 19:16

Test Message: No Sample Given

Lab #: 605902574 06/28/2016 00:56:33

0.099 00:58

Test Message: Deficient Sample

Lab #: 605902575 06/28/2016 03:00:49

0.099 03:03

0.096 03:09

Test Message: 0.21

Lab #: 605902576 06/29/2016 03:34:51

0.098 03:37

Test Message: Subject Test Refused

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 605902577 06/29/2016 03:42:24

0.098 03:44

0.097 03:49

Test Message: 0.18

Lab #: 605902578 06/29/2016 20:25:33

0.098 20:27

Test Message: Deficient Sample

Lab #: 605902579 06/29/2016 20:36:06

0.098 20:38

0.096 20:47

Test Message: 0.18

Lab #: 605902580 06/30/2016 02:26:27

0.098 02:28

0.098 02:33

Test Message: 0.00

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000718

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 071801689 05/07/2016 22:41:48

0.100 22:43

0.100 22:48

Test Message: 0.08

Lab #: 071801690 05/15/2016 02:11:14

0.100 02:14

Test Message: Sequence Aborted

Lab #: 071801691 05/21/2016 23:52:17

0.102 23:55

Test Message: Sequence Aborted

Lab #: 071801692 05/22/2016 23:18:32

0.105 23:20

0.101 23:25

Test Message: 0.17

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000726

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 072600327 05/14/2016 01:14:46

0.099 01:17

0.098 01:22

Test Message: 0.21

Lab #: 072600328 05/22/2016 00:38:19

0.098 00:43

0.098 00:48

Test Message: 0.11

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000728

GEBS Results

GEBS VALUE = 0.100

Lab #: 072802395 05/01/2016 00:01:18

0.101 00:03

0.100 00:08

Test Message: 0.18

Lab #: 072802396 05/08/2016 23:37:53

0.099 23:39

0.100 23:44

Test Message: 0.16

Lab #: 072802397 05/14/2016 22:46:05

0.100 22:47

0.099 22:53

Test Message: 0.09

Lab #: 072802398 05/17/2016 03:39:23

0.100 03:42

0.098 03:48

Test Message: 0.19

Lab #: 072802399 05/23/2016 03:26:53

0.101 03:28

Test Message: Sequence Aborted

Lab #: 072802400 05/23/2016 03:49:40

0.101 03:51

0.100 03:57

Test Message: 0.12

Lab #: 072802401 05/24/2016 21:36:50

0.101 21:39

0.100 21:46

Test Message: 0.21

Lab #: 072802402 05/29/2016 00:53:50

0.101 00:55

0.100 01:01

Test Message: 0.10



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500644 05/15/2016 00:31:29

0.099 00:33

0.099 00:38

Test Message: 0.05

Lab #: 605500645 05/15/2016 02:14:37

0.100 02:17

Test Message: No Sample Given

Lab #: 605500646 05/15/2016 02:22:40

0.100 02:25

0.099 02:31

Test Message: 0.03

Lab #: 605500647 05/21/2016 00:38:43

0.100 00:40

0.099 00:46

Test Message: 0.11

Lab #: 605500648 05/21/2016 02:26:03

0.100 02:28

0.100 02:36

Test Message: 0.10

Lab #: 605500649 05/21/2016 03:07:42

0.100 03:09

0.098 03:16

Test Message: 0.09

Lab #: 605500650 05/28/2016 00:42:41

0.100 00:45

0.099 00:50

Test Message: 0.10

Lab #: 605500651 05/28/2016 01:27:48

0.099 01:30

0.098 01:35

Test Message: 0.14

Lab #: 605500652 05/28/2016 02:59:19

0.099 03:01

Test Message: Deficient Sample

Lab #: 605500653 05/28/2016 03:06:07

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 605500653 05/28/2016 03:06:07

0.098 03:07

0.097 03:13

Test Message: 0.20

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006057

GEBS Results

GEBS VALUE = 0.100

Lab #: 605700500 05/15/2016 02:04:21

0.097 02:06

0.096 02:11

Test Message: 0.12

Lab #: 605700501 05/20/2016 23:38:17

0.097 23:58

Test Message: Sequence Aborted

Lab #: 605700502 05/21/2016 00:16:46

0.098 00:19

0.097 00:25

Test Message: 0.14

Lab #: 605700503 05/21/2016 01:13:53

0.098 01:16

0.096 01:23

Test Message: 0.12

Lab #: 605700504 05/21/2016 02:17:48

0.098 02:35

0.097 02:41

Test Message: 0.16

Lab #: 605700505 05/21/2016 03:25:28

0.098 03:28

0.098 03:33

Test Message: 0.12

Lab #: 605700506 05/22/2016 00:45:02

0.098 00:47

0.097 00:53

Test Message: 0.18

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801047 05/02/2016 18:31:00

0.100 18:34

0.097 18:39

Test Message: 0.21

Lab #: 605801048 05/04/2016 17:01:57

0.099 17:03

Test Message: Subject Test Refused

Lab #: 605801049 05/04/2016 17:10:09

0.101 17:12

0.098 17:19

Test Message: 0.20

Lab #: 605801050 05/05/2016 02:31:39

0.100 02:33

0.099 02:38

Test Message: 0.11

Lab #: 605801051 05/06/2016 03:27:28

0.099 03:29

0.097 03:34

Test Message: 0.18

Lab #: 605801052 05/12/2016 20:02:37

0.098 20:05

0.099 20:10

Test Message: 0.07

Lab #: 605801053 05/13/2016 08:25:22

0.100 08:27

0.098 08:33

Test Message: 0.07

Lab #: 605801054 05/14/2016 03:07:54

0.100 03:10

0.099 03:15

Test Message: 0.07

Lab #: 605801055 05/15/2016 05:17:10

0.100 05:19

0.099 05:23

Test Message: 0.19

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801056 05/17/2016 02:36:20

0.100 02:38

0.098 02:43

Test Message: 0.13

Lab #: 605801057 05/21/2016 00:31:27

0.100 00:33

0.099 00:39

Test Message: 0.18

Lab #: 605801058 05/23/2016 02:05:03

0.101 02:06

0.099 02:12

Test Message: 0.10

Lab #: 605801059 05/25/2016 06:00:11

0.100 06:02

0.099 06:07

Test Message: 0.10

Lab #: 605801060 05/26/2016 02:33:09

0.099 02:35

0.098 02:45

Test Message: 0.20

Lab #: 605801061 05/27/2016 02:33:07

0.098 02:36

0.098 02:41

Test Message: 0.19

Lab #: 605801062 05/28/2016 02:04:02

0.099 02:06

0.098 02:12

Test Message: 0.12

Lab #: 605801063 05/28/2016 23:38:24

0.101 23:40

0.100 23:45

Test Message: 0.09

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902431 05/01/2016 00:41:21

0.099 00:43

0.098 00:48

Test Message: 0.10

Lab #: 605902432 05/01/2016 03:29:58

0.100 03:32

0.100 03:37

Test Message: 0.07

Lab #: 605902433 05/01/2016 18:24:52

0.100 18:27

0.099 18:33

Test Message: 0.19

Lab #: 605902434 05/01/2016 22:56:22

0.100 22:58

0.098 23:04

Test Message: 0.19

Lab #: 605902435 05/02/2016 05:20:48

0.100 05:23

0.096 05:28

Test Message: 0.20

Lab #: 605902436 05/03/2016 22:31:04

0.100 22:33

0.099 22:39

Test Message: 0.19

Lab #: 605902437 05/04/2016 03:00:06

0.100 03:02

0.098 03:07

Test Message: 0.22

Lab #: 605902438 05/04/2016 16:55:23

0.100 17:14

Test Message: No Sample Given

Lab #: 605902439 05/05/2016 00:29:31

0.100 00:31

0.099 00:36

Test Message: 0.10

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902440 05/05/2016 23:45:33

0.100 23:47

0.099 23:52

Test Message: 0.08

Lab #: 605902441 05/06/2016 01:15:34

0.099 01:17

0.096 01:24

Test Message: 0.20

Lab #: 605902442 05/06/2016 03:01:01

0.099 03:04

0.099 03:09

Test Message: 0.09

Lab #: 605902443 05/07/2016 01:23:27

0.099 01:25

0.099 01:31

Test Message: 0.08

Lab #: 605902444 05/07/2016 03:34:59

0.100 03:37

0.100 03:42

Test Message: 0.09

Lab #: 605902445 05/07/2016 13:10:55

0.100 13:13

0.099 13:18

Test Message: 0.21

Lab #: 605902446 05/09/2016 10:49:22

0.101 10:52

0.099 10:57

Test Message: 0.13

Lab #: 605902447 05/10/2016 18:09:21

0.100 18:11

0.098 18:16

Test Message: 0.14

Lab #: 605902448 05/11/2016 01:25:17

0.099 01:28

0.096 01:33

Test Message: 0.25

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902449 05/12/2016 04:00:25

0.099 04:02

0.097 04:09

Test Message: 0.17

Lab #: 605902450 05/13/2016 01:52:48

0.099 01:55

0.098 02:00

Test Message: 0.15

Lab #: 605902451 05/13/2016 04:29:01

0.099 04:31

0.098 04:36

Test Message: 0.09

Lab #: 605902452 05/13/2016 07:59:54

0.100 08:02

Test Message: Deficient Sample

Lab #: 605902453 05/13/2016 08:07:00

0.099 08:09

Test Message: Subject Test Refused

Lab #: 605902454 05/13/2016 18:18:52

0.099 18:22

0.099 18:27

Test Message: 0.19

Lab #: 605902455 05/13/2016 23:19:26

0.097 23:21

0.098 23:29

Test Message: 0.16

Lab #: 605902456 05/14/2016 01:20:50

0.099 01:23

Test Message: No Sample Given

Lab #: 605902457 05/15/2016 04:22:08

0.099 04:24

0.098 04:29

Test Message: 0.16

Lab #: 605902458 05/16/2016 19:06:19

0.099 19:08

0.098 19:13



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

---

Test Message: 0.22

Lab #: 605902459 05/16/2016 22:19:03

0.099 22:21

0.098 22:26

Test Message: 0.15

Lab #: 605902460 05/17/2016 02:29:08

0.099 02:32

0.097 02:38

Test Message: 0.16

Lab #: 605902461 05/17/2016 10:32:07

0.099 10:35

Test Message: Deficient Sample

Lab #: 605902462 05/17/2016 10:40:06

0.099 10:42

Test Message: Subject Test Refused

Lab #: 605902463 05/17/2016 23:32:28

0.099 23:35

0.098 23:40

Test Message: 0.13

Lab #: 605902464 05/18/2016 02:34:47

0.099 02:37

0.098 02:42

Test Message: 0.06

Lab #: 605902465 05/19/2016 02:20:58

0.099 02:23

0.098 02:27

Test Message: 0.08

Lab #: 605902466 05/19/2016 09:22:42

0.100 09:25

Test Message: Deficient Sample

Lab #: 605902467 05/19/2016 09:38:35

0.099 09:40

0.097 09:47

Test Message: 0.14

Lab #: 605902468 05/20/2016 00:58:59

0.098 01:01

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902468 05/20/2016 00:58:59

0.098 01:06

Test Message: 0.13

Lab #: 605902469 05/20/2016 02:50:07

0.099 02:52

0.098 02:59

Test Message: 0.16

Lab #: 605902470 05/20/2016 23:16:20

0.099 23:19

0.098 23:24

Test Message: 0.17

Lab #: 605902471 05/21/2016 00:23:27

0.099 00:25

Test Message: Deficient Sample

Lab #: 605902472 05/21/2016 02:24:07

0.099 02:26

0.099 02:31

Test Message: 0.09

Lab #: 605902473 05/21/2016 06:05:11

0.100 06:07

0.100 06:13

Test Message: 0.02

Lab #: 605902474 05/21/2016 23:42:53

0.099 23:44

0.098 23:49

Test Message: 0.21

Lab #: 605902475 05/22/2016 03:37:13

0.098 03:40

0.095 03:45

Test Message: 0.14

Lab #: 605902476 05/22/2016 19:11:32

0.099 19:13

0.097 19:20

Test Message: 0.23

Lab #: 605902477 05/22/2016 20:25:30

0.099 20:27

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902477 05/22/2016 20:25:30

0.099 20:33

Test Message: 0.16

Lab #: 605902478 05/23/2016 00:55:05

0.100 01:05

Test Message: Sequence Aborted

Lab #: 605902479 05/24/2016 01:44:26

0.099 01:46

0.097 01:53

Test Message: 0.25

Lab #: 605902480 05/25/2016 04:54:30

0.098 04:57

0.098 05:02

Test Message: 0.14

Lab #: 605902481 05/26/2016 15:42:33

0.100 15:45

0.099 15:50

Test Message: 0.21

Lab #: 605902482 05/26/2016 16:01:10

0.100 16:03

0.098 16:08

Test Message: 0.22

Lab #: 605902483 05/27/2016 03:49:49

0.099 03:52

0.097 03:58

Test Message: 0.19

Lab #: 605902484 05/27/2016 20:34:06

0.098 20:36

0.099 20:41

Test Message: 0.25

Lab #: 605902485 05/28/2016 11:02:34

0.099 11:05

0.098 11:10

Test Message: 0.11

Lab #: 605902486 05/28/2016 17:49:12

0.100 17:51

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902486 05/28/2016 17:49:12

0.099 17:56

Test Message: 0.15

Lab #: 605902487 05/29/2016 00:15:43

0.099 00:17

0.098 00:22

Test Message: 0.07

Lab #: 605902488 05/29/2016 00:29:39

0.099 00:32

0.097 00:37

Test Message: 0.16

Lab #: 605902489 05/29/2016 03:06:00

0.098 03:08

0.098 03:13

Test Message: 0.16

Lab #: 605902490 05/29/2016 05:44:22

0.099 05:46

0.098 05:51

Test Message: 0.10

Lab #: 605902491 05/29/2016 12:03:23

0.099 12:05

Test Message: Subject Test Refused

Lab #: 605902492 05/29/2016 12:08:14

0.100 12:10

Test Message: Subject Test Refused

Lab #: 605902493 05/29/2016 12:25:29

0.100 12:27

Test Message: Subject Test Refused

Lab #: 605902494 05/30/2016 04:13:36

0.099 04:15

Test Message: Sequence Aborted

Lab #: 605902495 05/30/2016 14:47:51

0.099 14:49

Test Message: Wait Fifteen Minutes; Repeat Test

Lab #: 605902496 05/30/2016 14:52:23

0.099 14:54

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

---

Test Message: Deficient Sample

Lab #: 605902497 05/30/2016 23:53:35

0.099 23:57

0.098 00:04

Test Message: 0.13

Lab #: 605902498 05/31/2016 05:05:16

0.099 05:07

Test Message: Deficient Sample

Lab #: 605902499 05/31/2016 05:12:11

0.098 05:14

0.099 05:19

Test Message: 0.15

Lab #: 605902500 05/31/2016 18:00:02

0.098 18:02

Test Message: Radio Frequency Interferent Detected

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000718

GEBS Results

GEBS VALUE = 0.100

Lab #: 071801681 04/01/2016 19:53:56

0.101 19:55

0.100 20:00

Test Message: 0.08

Lab #: 071801682 04/02/2016 01:52:48

0.102 01:54

0.101 02:00

Test Message: 0.10

Lab #: 071801683 04/12/2016 02:01:08

0.101 02:02

0.099 02:09

Test Message: 0.20

Lab #: 071801684 04/16/2016 19:31:26

0.103 19:33

0.102 19:38

Test Message: 0.19

Lab #: 071801685 04/20/2016 00:51:04

0.101 00:52

0.100 00:59

Test Message: 0.16

Lab #: 071801686 04/21/2016 02:27:48

0.102 02:29

0.100 02:35

Test Message: 0.23

Lab #: 071801687 04/22/2016 20:28:09

0.100 20:30

0.100 20:34

Test Message: 0.19

Lab #: 071801688 04/28/2016 21:42:36

0.101 21:45

Test Message: Sequence Aborted

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000726

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 072600324 04/23/2016 23:30:18

0.099 23:32

0.098 23:37

Test Message: 0.13

Lab #: 072600325 04/28/2016 11:48:02

0.099 11:50

0.098 11:55

Test Message: 0.08

Lab #: 072600326 04/29/2016 23:53:56

0.100 23:56

0.098 00:01

Test Message: 0.19

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500621 04/01/2016 00:01:11

0.100 00:03

0.098 00:09

Test Message: 0.11

Lab #: 605500622 04/01/2016 00:51:40

0.099 00:53

Test Message: Deficient Sample

Lab #: 605500623 04/01/2016 00:58:52

0.098 01:00

0.098 01:05

Test Message: 0.20

Lab #: 605500624 04/01/2016 01:09:49

0.099 01:11

0.097 01:16

Test Message: 0.23

Lab #: 605500625 04/01/2016 01:56:41

0.098 01:58

0.098 02:03

Test Message: 0.06

Lab #: 605500626 04/01/2016 02:26:44

0.099 02:28

0.097 02:34

Test Message: 0.15

Lab #: 605500627 04/01/2016 02:45:13

0.098 02:47

0.098 02:52

Test Message: 0.19

Lab #: 605500628 04/02/2016 23:39:52

0.099 23:42

Test Message: Deficient Sample

Lab #: 605500629 04/02/2016 23:47:26

0.098 23:49

0.098 23:54

Test Message: 0.13

Lab #: 605500630 04/03/2016 00:21:59



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500630 04/03/2016 00:21:59

0.100 00:24

0.097 00:30

Test Message: 0.13

Lab #: 605500631 04/03/2016 00:50:15

0.099 00:52

0.098 00:56

Test Message: 0.03

Lab #: 605500632 04/03/2016 01:00:38

0.099 01:02

0.098 01:08

Test Message: 0.12

Lab #: 605500633 04/03/2016 01:31:08

0.099 01:34

0.098 01:40

Test Message: 0.18

Lab #: 605500634 04/03/2016 01:51:05

0.100 01:53

0.098 01:57

Test Message: 0.11

Lab #: 605500636 04/16/2016 00:24:48

0.101 00:26

0.100 00:32

Test Message: 0.10

Lab #: 605500637 04/16/2016 00:35:43

0.099 00:37

0.099 00:42

Test Message: 0.06

Lab #: 605500638 04/16/2016 01:02:21

0.100 01:04

0.099 01:09

Test Message: 0.11

Lab #: 605500639 04/16/2016 01:32:39

0.101 01:35

0.099 01:39

Test Message: 0.17

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 605500640 04/29/2016 23:51:58

0.100 23:55

0.099 00:00

Test Message: 0.09

Lab #: 605500641 04/30/2016 00:35:49

0.098 00:37

0.098 00:43

Test Message: 0.12

Lab #: 605500642 04/30/2016 02:09:09

0.097 02:12

0.095 02:18

Test Message: 0.14

Lab #: 605500643 04/30/2016 02:28:40

0.097 02:30

0.098 02:36

Test Message: 0.24

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006057

GEBS Results

GEBS VALUE = 0.100

Lab #: 605700479 04/01/2016 00:32:36

0.099 00:35

0.096 00:40

Test Message: 0.17

Lab #: 605700480 04/01/2016 01:08:53

0.098 01:12

0.097 01:18

Test Message: 0.21

Lab #: 605700482 04/01/2016 01:40:03

0.099 01:42

Test Message: Subject Test Refused

Lab #: 605700483 04/01/2016 01:44:11

0.097 01:46

Test Message: Sequence Aborted

Lab #: 605700484 04/01/2016 01:50:29

0.098 01:52

0.098 01:57

Test Message: 0.08

Lab #: 605700485 04/01/2016 02:25:48

0.098 02:27

Test Message: Radio Frequency Interferent Detected

Lab #: 605700486 04/01/2016 02:28:38

0.098 02:30

Test Message: Deficient Sample

Lab #: 605700488 04/03/2016 00:00:35

0.097 00:02

0.096 00:10

Test Message: 0.13

Lab #: 605700489 04/03/2016 00:50:36

0.098 00:52

0.096 00:57

Test Message: 0.08

Lab #: 605700490 04/03/2016 01:27:02

0.098 01:29

0.097 01:34

Test Message: 0.15

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006057

GEBS Results

GEBS VALUE = 0.100

Lab #: 605700491 04/03/2016 02:01:42

0.096 02:03

0.098 02:08

Test Message: 0.11

Lab #: 605700492 04/03/2016 02:18:47

0.098 02:20

0.097 02:26

Test Message: 0.09

Lab #: 605700493 04/16/2016 00:33:45

0.099 00:35

0.099 00:41

Test Message: 0.02

Lab #: 605700494 04/16/2016 01:38:48

0.099 01:40

0.097 01:45

Test Message: 0.07

Lab #: 605700495 04/16/2016 02:02:14

0.099 02:04

0.098 02:09

Test Message: 0.01

Lab #: 605700496 04/16/2016 02:33:18

0.099 02:35

0.097 02:41

Test Message: 0.17

Lab #: 605700497 04/30/2016 00:36:58

0.097 00:39

0.096 00:44

Test Message: 0.15

Lab #: 605700498 04/30/2016 02:10:01

0.096 02:12

0.094 02:17

Test Message: 0.15

Lab #: 605700499 04/30/2016 02:37:56

0.097 02:40

0.096 02:46

Test Message: 0.08

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801032 04/01/2016 03:10:07

0.099 03:12

0.098 03:17

Test Message: 0.13

Lab #: 605801033 04/03/2016 14:33:29

0.099 14:35

0.096 14:40

Test Message: 0.18

Lab #: 605801034 04/04/2016 21:11:59

0.100 21:14

0.099 21:19

Test Message: 0.09

Lab #: 605801035 04/06/2016 00:17:26

0.099 00:19

Test Message: Deficient Sample

Lab #: 605801036 04/09/2016 04:01:09

0.099 04:03

0.098 04:10

Test Message: 0.11

Lab #: 605801037 04/15/2016 23:28:11

0.100 23:30

0.100 23:36

Test Message: 0.06

Lab #: 605801038 04/19/2016 04:26:31

0.101 04:28

Test Message: Sequence Aborted

Lab #: 605801039 04/23/2016 03:45:49

0.098 03:48

0.096 03:53

Test Message: 0.20

Lab #: 605801040 04/27/2016 13:44:15

0.099 13:47

0.099 13:51

Test Message: 0.00

Lab #: 605801041 04/27/2016 13:55:44

*Intox training class on 4/27/16  
by 5/2/16*

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801041 04/27/2016 13:55:44

0.099 13:58

Test Message: Subject Test Refused

Lab #: 605801044 04/27/2016 14:07:39

0.100 14:09

Test Message: Wait Fifteen Minutes; Repeat Test

Lab #: 605801045 04/27/2016 19:14:25

0.100 19:16

0.099 19:21

Test Message: 0.00

Lab #: 605801046 04/30/2016 03:36:38

0.100 03:38

0.098 03:44

Test Message: 0.15

*Intox training class on 4/27/16  
jt 5/2/16*

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902342 04/01/2016 00:09:41

0.100 00:11

0.098 00:18

Test Message: 0.15

Lab #: 605902343 04/01/2016 01:01:55

0.099 01:06

0.099 01:11

Test Message: 0.12

Lab #: 605902344 04/01/2016 02:44:51

0.099 02:47

0.099 02:52

Test Message: 0.05

Lab #: 605902345 04/01/2016 04:02:31

0.099 04:04

Test Message: Deficient Sample

Lab #: 605902346 04/01/2016 04:09:17

0.099 04:11

0.098 04:16

Test Message: 0.20

Lab #: 605902347 04/01/2016 04:42:37

0.099 04:44

0.098 04:49

Test Message: 0.14

Lab #: 605902348 04/01/2016 20:32:48

0.099 20:35

Test Message: Subject Test Refused

Lab #: 605902349 04/01/2016 21:47:53

0.100 21:49

0.099 21:54

Test Message: 0.17

Lab #: 605902350 04/02/2016 02:34:15

0.099 02:36

0.099 02:41

Test Message: 0.09

Lab #: 605902351 04/02/2016 03:51:49

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902351 04/02/2016 03:51:49

0.099 03:53

0.099 03:58

Test Message: 0.08

Lab #: 605902352 04/02/2016 04:58:12

0.099 05:00

0.098 05:05

Test Message: 0.12

Lab #: 605902353 04/02/2016 08:50:13

0.100 08:52

0.099 08:57

Test Message: 0.10

Lab #: 605902354 04/02/2016 16:23:10

0.100 16:25

0.098 16:31

Test Message: 0.15

Lab #: 605902355 04/03/2016 02:07:19

0.099 02:09

Test Message: Deficient Sample

Lab #: 605902356 04/03/2016 06:01:23

0.099 06:03

0.098 06:09

Test Message: 0.16

Lab #: 605902357 04/03/2016 22:28:56

0.099 22:30

0.098 22:35

Test Message: 0.08

Lab #: 605902358 04/04/2016 06:25:58

0.099 06:28

0.099 06:32

Test Message: 0.12

Lab #: 605902359 04/05/2016 01:54:54

0.099 01:56

0.098 02:01

Test Message: 0.12

Lab #: 605902360 04/06/2016 01:18:34



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902360 04/06/2016 01:18:34

0.098 01:20

0.096 01:27

Test Message: 0.12

Lab #: 605902361 04/07/2016 02:06:47

0.099 02:09

0.096 02:17

Test Message: 0.19

Lab #: 605902362 04/07/2016 13:09:11

0.099 13:19

0.099 13:25

Test Message: 0.02

Lab #: 605902363 04/08/2016 00:30:54

0.100 00:33

0.098 00:38

Test Message: 0.20

Lab #: 605902364 04/08/2016 03:04:00

0.100 03:05

0.098 03:11

Test Message: 0.16

Lab #: 605902365 04/08/2016 22:01:13

0.099 22:03

0.100 22:08

Test Message: 0.19

Lab #: 605902366 04/09/2016 02:10:52

0.099 02:12

0.099 02:18

Test Message: 0.20

Lab #: 605902367 04/09/2016 03:56:25

0.099 03:59

0.099 04:04

Test Message: 0.20

Lab #: 605902368 04/09/2016 06:05:40

0.098 06:07

0.098 06:12

Test Message: 0.21

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902369 04/10/2016 07:45:26

0.099 07:47

Test Message: Subject Test Refused

Lab #: 605902370 04/11/2016 00:27:40

0.100 00:29

0.097 00:35

Test Message: 0.17

Lab #: 605902371 04/11/2016 03:36:13

0.100 03:38

0.099 03:43

Test Message: 0.08

Lab #: 605902372 04/12/2016 03:43:36

0.099 03:46

Test Message: Deficient Sample

Lab #: 605902373 04/12/2016 03:51:04

0.099 03:53

0.098 04:01

Test Message: 0.15

Lab #: 605902374 04/12/2016 04:07:37

0.098 04:10

0.097 04:15

Test Message: 0.17

Lab #: 605902375 04/13/2016 00:22:49

0.100 00:24

0.098 00:29

Test Message: 0.20

Lab #: 605902376 04/13/2016 19:13:58

0.100 19:16

0.098 19:20

Test Message: 0.17

Lab #: 605902377 04/13/2016 23:02:00

0.099 23:03

0.098 23:08

Test Message: 0.10

Lab #: 605902379 04/13/2016 23:36:14

0.099 23:38

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902379 04/13/2016 23:36:14

0.097 23:43

Test Message: 0.32

Lab #: 605902380 04/14/2016 00:01:49

0.099 00:03

Test Message: Sequence Aborted

Lab #: 605902381 04/14/2016 00:18:32

0.098 00:20

0.096 00:28

Test Message: 0.21

Lab #: 605902382 04/14/2016 01:27:28

0.100 01:30

0.098 01:36

Test Message: 0.22

Lab #: 605902383 04/14/2016 15:09:56

0.099 15:12

0.098 15:17

Test Message: 0.17

Lab #: 605902384 04/14/2016 20:37:29

0.099 20:40

0.098 20:44

Test Message: 0.13

Lab #: 605902385 04/15/2016 03:04:22

0.100 03:07

0.095 03:12

Test Message: 0.11

Lab #: 605902386 04/15/2016 16:44:49

0.099 16:46

0.097 16:55

Test Message: 0.32

Lab #: 605902387 04/15/2016 22:06:02

0.100 22:12

Test Message: No Sample Given

Lab #: 605902388 04/15/2016 23:08:55

0.101 23:11

0.099 23:16

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Test Message: 0.17

Lab #: 605902389 04/16/2016 02:40:06

0.100 02:41

0.099 02:47

Test Message: 0.11

Lab #: 605902390 04/16/2016 03:12:58

0.099 03:15

0.099 03:22

Test Message: 0.24

Lab #: 605902391 04/16/2016 04:21:58

0.101 04:24

0.099 04:30

Test Message: 0.10

Lab #: 605902392 04/16/2016 18:45:16

0.100 18:47

Test Message: Sequence Aborted

Lab #: 605902393 04/16/2016 20:56:42

0.101 20:59

0.098 21:04

Test Message: 0.18

Lab #: 605902394 04/17/2016 04:01:47

0.101 04:04

0.099 04:09

Test Message: 0.15

Lab #: 605902395 04/17/2016 08:01:34

0.100 08:04

0.100 08:08

Test Message: 0.10

Lab #: 605902396 04/18/2016 01:28:53

0.101 01:31

0.099 01:36

Test Message: 0.22

Lab #: 605902397 04/18/2016 05:42:37

0.100 05:53

0.100 05:59

Test Message: 0.14

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902398 04/19/2016 04:32:42

0.100 04:34

0.099 04:41

Test Message: 0.11

Lab #: 605902399 04/19/2016 05:01:08

0.100 05:05

0.099 05:11

Test Message: 0.24

Lab #: 605902400 04/19/2016 21:07:39

0.100 21:09

Test Message: Sequence Aborted

Lab #: 605902401 04/20/2016 01:58:30

0.100 02:00

0.100 02:05

Test Message: 0.09

Lab #: 605902402 04/22/2016 03:19:56

0.100 03:22

Test Message: Improper Sample

Lab #: 605902403 04/22/2016 03:23:52

0.099 03:26

0.097 03:30

Test Message: 0.18

Lab #: 605902404 04/23/2016 03:30:17

0.100 03:32

0.099 03:37

Test Message: 0.11

Lab #: 605902405 04/23/2016 22:07:35

0.100 22:09

0.098 22:14

Test Message: 0.18

Lab #: 605902406 04/23/2016 23:07:40

0.100 23:09

0.099 23:15

Test Message: 0.18

Lab #: 605902407 04/24/2016 00:26:25

0.100 00:28

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Test Message: Deficient Sample

Lab #: 605902408 04/24/2016 00:33:37

0.098 00:35

Test Message: Deficient Sample

Lab #: 605902409 04/24/2016 00:41:55

0.098 00:44

Test Message: Deficient Sample

Lab #: 605902410 04/25/2016 00:03:35

0.098 00:05

0.099 00:10

Test Message: 0.17

Lab #: 605902411 04/25/2016 03:17:40

0.100 03:20

0.097 03:25

Test Message: 0.28

Lab #: 605902412 04/26/2016 00:10:10

0.101 00:12

0.098 00:17

Test Message: 0.18

Lab #: 605902413 04/26/2016 00:53:56

0.100 00:55

0.099 01:01

Test Message: 0.16

Lab #: 605902414 04/26/2016 02:33:29

0.100 02:35

0.099 02:41

Test Message: 0.09

Lab #: 605902415 04/26/2016 02:48:49

0.100 02:51

Test Message: Deficient Sample

Lab #: 605902416 04/26/2016 02:58:51

0.099 03:00

0.100 03:06

Test Message: 0.11

Lab #: 605902417 04/26/2016 23:59:44

0.100 00:01

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902417 04/26/2016 23:59:44

0.098 00:07

Test Message: 0.13

Lab #: 605902418 04/27/2016 13:02:59

0.100 13:05

0.098 13:10

Test Message: 0.00

Lab #: 605902419 04/27/2016 13:15:15

0.099 13:18

0.097 13:22

Test Message: 0.00

Lab #: 605902420 04/27/2016 16:53:40

0.099 16:55

0.097 17:03

Test Message: 0.32

Lab #: 605902421 04/27/2016 17:27:41

0.100 17:30

0.099 17:36

Test Message: 0.33

Lab #: 605902422 04/27/2016 18:53:13

0.100 18:56

Test Message: Sequence Aborted

Lab #: 605902423 04/27/2016 19:02:55

0.098 19:05

Test Message: Deficient Sample

Lab #: 605902424 04/28/2016 02:15:12

0.100 02:18

0.099 02:23

Test Message: 0.17

Lab #: 605902425 04/30/2016 02:19:02

0.100 02:21

0.099 02:26

Test Message: 0.25

Lab #: 605902426 04/30/2016 03:07:17

0.100 03:09

Test Message: Deficient Sample

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 605902427 04/30/2016 03:16:58

0.100 03:18

Test Message: Deficient Sample

Lab #: 605902428 04/30/2016 03:24:57

0.098 03:26

Test Message: Improper Sample

Lab #: 605902429 04/30/2016 03:28:55

0.099 03:30

0.097 03:36

Test Message: 0.27

Lab #: 605902430 04/30/2016 04:34:50

0.099 04:37

0.099 04:44

Test Message: 0.10



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000718

GEBS Results

GEBS VALUE = 0.100

Lab #: 071801666 03/06/2016 03:09:38

0.101 03:11

0.100 03:16

Test Message: 0.20

Lab #: 071801667 03/08/2016 23:05:40

0.102 23:07

0.097 23:14

Test Message: 0.28

Lab #: 071801668 03/09/2016 23:09:36

0.102 23:11

0.098 23:19

Test Message: 0.17

Lab #: 071801669 03/12/2016 19:32:36

0.101 19:35

Test Message: Wait Fifteen Minutes; Repeat Test

Lab #: 071801670 03/12/2016 19:53:52

0.102 19:55

0.099 20:00

Test Message: 0.12

Lab #: 071801671 03/17/2016 22:40:56

0.101 22:42

0.098 22:50

Test Message: Radio Frequency Interferent Detected

Lab #: 071801672 03/18/2016 01:28:52

0.101 01:31

Test Message: Sequence Aborted

Lab #: 071801673 03/18/2016 04:15:43

0.102 04:18

0.100 04:26

Test Message: 0.13

Lab #: 071801674 03/19/2016 00:25:27

0.101 00:27

0.101 00:32

Test Message: 0.10

Lab #: 071801675 03/23/2016 02:04:35