

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000719

GEBS Results

GEBS VALUE = 0.100

Lab #: 071904259 03/04/2017 02:27:31

0.099 02:29

0.099 02:34

Test Message: 0.17

Lab #: 071904260 03/05/2017 00:04:21

0.099 00:06

0.098 00:10

Test Message: 0.22

Lab #: 071904261 03/10/2017 20:32:12

0.099 20:34

0.098 20:38

Test Message: 0.18

Lab #: 071904262 03/13/2017 22:36:55

0.099 22:39

0.097 22:44

Test Message: 0.19

Lab #: 071904263 03/18/2017 01:01:39

0.099 01:03

0.099 01:08

Test Message: 0.17

Lab #: 071904264 03/19/2017 21:32:19

0.099 21:36

0.099 21:40

Test Message: 0.08

Lab #: 071904265 03/20/2017 00:07:20

0.100 00:09

0.099 00:13

Test Message: 0.19

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000726

GEBS Results

GEBS VALUE = 0.100

Lab #: 072600378 03/02/2017 23:36:10

0.100 23:38

0.100 23:43

Test Message: 0.11

Lab #: 072600379 03/03/2017 23:10:19

0.099 23:12

0.098 23:20

Test Message: 0.10

Lab #: 072600380 03/09/2017 22:40:37

0.100 22:42

0.098 22:47

Test Message: 0.18

Lab #: 072600381 03/10/2017 15:57:15

0.100 15:59

0.099 16:04

Test Message: 0.19

Lab #: 072600382 03/25/2017 01:38:07

0.099 01:40

0.099 01:46

Test Message: 0.11

Lab #: 072600383 03/31/2017 21:38:04

0.100 21:40

0.100 21:45

Test Message: 0.20

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

Lab #: 072703653 03/03/2017 21:32:15

0.098 21:36

0.094 21:41

Test Message: 0.09

Lab #: 072703654 03/03/2017 23:13:42

0.097 23:16

0.095 23:21

Test Message: 0.08

Lab #: 072703655 03/04/2017 01:02:28

0.098 01:05

0.097 01:10

Test Message: 0.07

Lab #: 072703656 03/05/2017 02:58:28

0.099 03:00

Test Message: Deficient Sample

Lab #: 072703657 03/05/2017 03:05:15

0.098 03:07

0.096 03:15

Test Message: 0.15

Lab #: 072703658 03/05/2017 04:47:19

0.098 04:49

0.097 04:54

Test Message: 0.16

Lab #: 072703659 03/09/2017 23:54:44

0.097 23:57

0.097 00:02

Test Message: 0.12

Lab #: 072703660 03/10/2017 05:54:25

0.098 05:57

0.097 06:02

Test Message: 0.07

Lab #: 072703661 03/12/2017 04:46:18

0.098 04:48

0.096 04:54

Test Message: 0.16

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 072703662 03/16/2017 21:25:29

0.097 21:27

0.097 21:32

Test Message: 0.17

Lab #: 072703663 03/16/2017 23:06:45

0.098 23:09

0.097 23:15

Test Message: 0.03

Lab #: 072703666 03/30/2017 02:45:04

0.098 02:47

0.096 02:52

Test Message: 0.19

Lab #: 072703667 03/31/2017 03:59:54

0.098 04:03

0.096 04:08

Test Message: 0.16

Lab #: 072703668 03/31/2017 05:25:44

0.097 05:29

0.097 05:35

Test Message: 0.07

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000728

GEBS Results

GEBS VALUE = 0.100

Lab #: 072802475 03/02/2017 01:12:55

0.099 01:15

Test Message: Radio Frequency Interferent Detected

Lab #: 072802476 03/04/2017 02:59:16

0.101 03:01

0.099 03:06

Test Message: 0.11

Lab #: 072802477 03/12/2017 00:53:23

0.098 00:55

0.097 01:01

Test Message: Diagnostic Failed

Lab #: 072802478 03/21/2017 23:32:03

0.097 23:34

0.097 23:40

Test Message: 0.19

Lab #: 072802479 03/29/2017 23:41:08

0.099 23:43

0.098 23:48

Test Message: 0.13

Lab #: 072802480 03/30/2017 02:50:42

0.100 03:06

0.098 03:14

Test Message: 0.13

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500769 03/01/2017 00:13:42

0.099 00:15

0.099 00:21

Test Message: 0.09

Lab #: 605500770 03/01/2017 01:30:59

0.099 01:33

0.098 01:38

Test Message: 0.14

Lab #: 605500771 03/01/2017 02:09:00

0.099 02:11

0.098 02:17

Test Message: 0.15

Lab #: 605500772 03/01/2017 02:23:39

0.099 02:25

Test Message: Radio Frequency Interferent Detected

Lab #: 605500773 03/01/2017 02:27:49

0.099 02:29

0.098 02:34

Test Message: 0.07

Lab #: 605500774 03/01/2017 02:54:41

0.099 02:56

0.098 03:02

Test Message: 0.16

Lab #: 605500775 03/05/2017 02:22:27

0.099 02:25

0.098 02:31

Test Message: 0.16

Lab #: 605500776 03/10/2017 22:21:23

0.100 00:29

0.098 00:35

Test Message: 0.09

Lab #: 605500777 03/11/2017 00:50:05

0.100 00:52

0.097 00:59

Test Message: 0.11

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500778 03/11/2017 01:43:00

0.099 01:45

0.099 01:51

Test Message: 0.12

Lab #: 605500779 03/11/2017 02:48:04

0.099 02:50

0.095 02:57

Test Message: 0.17

Lab #: 605500780 03/18/2017 00:49:33

0.098 00:51

Test Message: Improper Sample

Lab #: 605500781 03/18/2017 00:52:23

0.098 00:54

0.097 00:59

Test Message: 0.12

Lab #: 605500782 03/18/2017 02:04:21

0.098 02:06

0.096 02:11

Test Message: 0.19

Lab #: 605500783 03/25/2017 23:26:35

0.100 23:29

0.100 23:34

Test Message: 0.08

Lab #: 605500784 03/26/2017 00:45:19

0.099 00:47

0.099 00:52

Test Message: 0.04

Lab #: 605500785 03/26/2017 00:54:47

0.099 00:57

0.098 01:02

Test Message: 0.03

Lab #: 605500786 03/26/2017 01:40:32

0.100 01:43

0.099 01:48

Test Message: 0.10

Lab #: 605500787 03/26/2017 01:56:59

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 605500787 03/26/2017 01:56:59

0.100 01:58

0.098 02:05

Test Message: 0.12

Lab #: 605500788 03/26/2017 02:10:49

0.099 02:13

0.098 02:18

Test Message: 0.13



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801648 03/01/2017 00:57:44

0.100 00:59

0.097 01:05

Test Message: 0.26

Lab #: 605801649 03/01/2017 03:17:52

0.100 03:19

Test Message: Deficient Sample

Lab #: 605801650 03/01/2017 03:28:41

0.100 03:30

0.097 03:36

Test Message: 0.19

Lab #: 605801651 03/01/2017 09:38:41

0.100 09:40

Test Message: Radio Frequency Interferent Detected

Lab #: 605801652 03/01/2017 09:44:36

0.099 09:46

Test Message: Deficient Sample

Lab #: 605801653 03/01/2017 09:56:30

0.098 09:58

0.095 10:05

Test Message: 0.10

Lab #: 605801654 03/02/2017 01:16:44

0.099 01:18

0.099 01:23

Test Message: 0.09

Lab #: 605801655 03/02/2017 23:30:25

0.099 23:32

0.099 23:38

Test Message: 0.22

Lab #: 605801656 03/03/2017 20:08:14

0.100 20:11

Test Message: No Sample Given

Lab #: 605801657 03/03/2017 20:20:45

0.099 20:23

0.099 20:28

Test Message: 0.14

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801658 03/03/2017 20:36:11

0.097 20:39

0.094 20:46

Test Message: 0.31

Lab #: 605801659 03/03/2017 21:23:08

0.099 21:25

Test Message: Deficient Sample

Lab #: 605801660 03/03/2017 21:29:25

0.096 21:31

Test Message: No Sample Given

Lab #: 605801661 03/03/2017 23:15:27

0.100 23:17

0.098 23:22

Test Message: 0.14

Lab #: 605801662 03/04/2017 00:58:26

0.100 01:01

0.098 01:06

Test Message: 0.12

Lab #: 605801663 03/04/2017 04:47:05

0.100 04:49

0.097 04:54

Test Message: 0.17

Lab #: 605801664 03/04/2017 19:28:34

0.099 19:31

Test Message: Improper Sample

Lab #: 605801665 03/05/2017 04:42:17

0.100 04:44

Test Message: Deficient Sample

Lab #: 605801666 03/05/2017 04:48:55

0.099 04:50

0.097 04:56

Test Message: 0.18

Lab #: 605801667 03/05/2017 12:09:07

0.099 12:12

0.097 12:18

Test Message: 0.21

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801668 03/05/2017 14:10:24

0.099 14:12

0.096 14:19

Test Message: 0.32

Lab #: 605801669 03/05/2017 15:43:33

0.099 15:46

0.098 15:51

Test Message: 0.17

Lab #: 605801670 03/06/2017 01:29:44

0.099 01:32

Test Message: Deficient Sample

Lab #: 605801671 03/06/2017 01:49:25

0.099 02:06

0.097 02:11

Test Message: 0.22

Lab #: 605801672 03/06/2017 17:27:34

0.100 17:29

0.098 17:34

Test Message: 0.23

Lab #: 605801673 03/06/2017 18:42:42

0.100 18:44

Test Message: Deficient Sample

Lab #: 605801674 03/06/2017 18:50:36

0.100 18:52

Test Message: Deficient Sample

Lab #: 605801675 03/06/2017 23:38:25

0.099 23:42

0.097 23:47

Test Message: 0.18

Lab #: 605801676 03/07/2017 03:34:41

0.100 03:37

0.097 03:44

Test Message: 0.23

Lab #: 605801677 03/07/2017 04:51:22

0.100 04:54

0.098 05:01

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Test Message: 0.23

Lab #: 605801678 03/08/2017 23:22:44

0.098 23:25

0.095 23:32

Test Message: 0.07

Lab #: 605801679 03/08/2017 23:35:49

0.099 23:38

Test Message: Subject Test Refused

Lab #: 605801680 03/08/2017 23:42:55

0.099 23:45

0.098 23:50

Test Message: 0.15

Lab #: 605801681 03/09/2017 04:54:35

0.099 04:57

Test Message: Subject Test Refused

Lab #: 605801682 03/09/2017 04:59:41

0.099 05:01

0.099 05:07

Test Message: 0.07

Lab #: 605801683 03/09/2017 05:09:30

0.099 05:11

Test Message: Sequence Aborted

Lab #: 605801684 03/09/2017 05:14:15

0.099 05:16

0.098 05:21

Test Message: 0.06

Lab #: 605801685 03/09/2017 22:35:44

0.099 22:38

0.097 22:43

Test Message: 0.10

Lab #: 605801686 03/09/2017 23:39:30

0.099 23:41

0.098 23:46

Test Message: 0.16

Lab #: 605801687 03/10/2017 01:13:15

0.099 01:15

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801687 03/10/2017 01:13:15

0.099 01:20

Test Message: 0.12

Lab #: 605801689 03/10/2017 18:21:45

0.099 18:23

Test Message: Radio Frequency Interferent Detected

Lab #: 605801690 03/10/2017 18:25:18

0.100 18:27

0.098 18:32

Test Message: 0.10

Lab #: 605801691 03/11/2017 01:59:21

0.100 02:01

Test Message: Deficient Sample

Lab #: 605801692 03/11/2017 08:33:17

0.099 08:37

0.098 08:42

Test Message: 0.12

Lab #: 605801693 03/12/2017 01:07:24

0.100 01:09

0.098 01:15

Test Message: 0.08

Lab #: 605801694 03/12/2017 03:35:28

0.099 03:37

0.099 03:42

Test Message: 0.14

Lab #: 605801695 03/12/2017 05:12:40

0.100 05:14

0.096 05:22

Test Message: 0.25

Lab #: 605801696 03/14/2017 00:54:42

0.099 00:56

0.098 01:02

Test Message: 0.09

Lab #: 605801697 03/14/2017 02:40:19

0.099 02:43

0.095 02:50

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Test Message: 0.21

Lab #: 605801698 03/15/2017 00:55:57

0.099 00:58

0.098 01:03

Test Message: 0.15

Lab #: 605801699 03/15/2017 02:28:00

0.099 02:30

0.096 02:35

Test Message: 0.17

Lab #: 605801700 03/15/2017 12:21:47

0.098 12:24

0.096 12:29

Test Message: 0.14

Lab #: 605801701 03/15/2017 23:02:28

0.099 23:19

0.098 23:24

Test Message: 0.16

Lab #: 605801702 03/17/2017 17:23:07

0.099 17:26

0.098 17:31

Test Message: 0.14

Lab #: 605801703 03/17/2017 22:21:50

0.099 22:24

0.097 22:29

Test Message: 0.23

Lab #: 605801704 03/18/2017 02:18:29

0.099 02:21

Test Message: Deficient Sample

Lab #: 605801705 03/18/2017 02:28:03

0.098 02:30

Test Message: Deficient Sample

Lab #: 605801706 03/18/2017 02:40:12

0.098 02:42

Test Message: Deficient Sample

Lab #: 605801707 03/18/2017 03:42:02

0.098 03:44

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801707 03/18/2017 03:42:02

0.097 03:49

Test Message: 0.14

Lab #: 605801708 03/18/2017 17:32:35

0.099 17:34

0.096 17:44

Test Message: 0.21

Lab #: 605801709 03/19/2017 00:27:51

0.098 00:29

0.097 00:35

Test Message: 0.12

Lab #: 605801710 03/19/2017 00:44:14

0.099 00:47

0.097 00:53

Test Message: 0.25

Lab #: 605801711 03/19/2017 03:33:54

0.098 03:36

0.096 03:42

Test Message: 0.19

Lab #: 605801712 03/19/2017 04:35:19

0.099 04:37

0.097 04:44

Test Message: 0.15

Lab #: 605801713 03/19/2017 08:32:55

0.099 08:34

0.097 08:40

Test Message: 0.10

Lab #: 605801714 03/19/2017 22:52:24

0.099 22:55

0.097 23:00

Test Message: 0.15

Lab #: 605801715 03/20/2017 00:07:24

0.099 00:10

0.097 00:15

Test Message: 0.13

Lab #: 605801716 03/20/2017 03:43:33

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801716 03/20/2017 03:43:33

0.099 03:46

0.096 03:51

Test Message: 0.19

Lab #: 605801717 03/20/2017 19:53:44

0.099 19:57

0.097 20:02

Test Message: 0.27

Lab #: 605801718 03/21/2017 00:34:50

0.099 00:37

0.098 00:42

Test Message: 0.13

Lab #: 605801719 03/21/2017 02:57:52

0.099 03:00

0.097 03:05

Test Message: 0.14

Lab #: 605801720 03/21/2017 11:14:58

0.100 11:16

0.097 11:21

Test Message: 0.30

Lab #: 605801721 03/23/2017 01:44:29

0.100 01:46

0.097 01:51

Test Message: 0.23

Lab #: 605801722 03/23/2017 15:12:33

0.100 15:14

0.099 15:20

Test Message: 0.14

Lab #: 605801723 03/23/2017 23:16:26

0.100 23:19

0.098 23:24

Test Message: 0.19

Lab #: 605801724 03/24/2017 01:48:25

0.100 01:51

0.098 01:56

Test Message: 0.25



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801725 03/24/2017 02:45:19

0.099 02:48

0.098 02:52

Test Message: 0.25

Lab #: 605801726 03/24/2017 22:24:03

0.099 22:26

Test Message: Subject Test Refused

Lab #: 605801727 03/25/2017 02:52:12

0.099 02:54

0.099 02:59

Test Message: 0.07

Lab #: 605801728 03/25/2017 03:19:49

0.099 03:23

0.095 03:28

Test Message: 0.22

Lab #: 605801729 03/25/2017 09:20:36

0.100 09:24

0.097 09:30

Test Message: 0.16

Lab #: 605801730 03/25/2017 10:42:36

0.099 10:45

0.097 10:50

Test Message: 0.26

Lab #: 605801731 03/26/2017 01:04:39

0.099 01:06

0.099 01:12

Test Message: 0.10

Lab #: 605801732 03/26/2017 02:09:44

0.100 02:12

0.097 02:17

Test Message: 0.19

Lab #: 605801733 03/26/2017 02:56:10

0.099 02:58

0.097 03:04

Test Message: 0.22

Lab #: 605801734 03/26/2017 21:12:27

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801734 03/26/2017 21:12:27

0.099 21:15

0.098 21:21

Test Message: 0.25

Lab #: 605801735 03/26/2017 21:28:24

0.099 22:02

Test Message: Deficient Sample

Lab #: 605801736 03/26/2017 22:07:45

0.099 22:09

Test Message: Sequence Aborted

Lab #: 605801737 03/27/2017 17:07:10

0.099 17:09

0.099 17:14

Test Message: 0.24

Lab #: 605801738 03/27/2017 19:55:31

0.101 19:57

Test Message: Radio Frequency Interferent Detected

Lab #: 605801739 03/27/2017 20:01:23

0.100 20:03

0.098 20:08

Test Message: 0.16

Lab #: 605801740 03/27/2017 22:24:44

0.100 22:27

0.095 22:33

Test Message: 0.20

Lab #: 605801741 03/28/2017 04:12:26

0.101 04:14

0.101 04:19

Test Message: 0.00

Lab #: 605801742 03/30/2017 02:18:25

0.100 02:21

0.099 02:26

Test Message: 0.12

Lab #: 605801743 03/30/2017 02:36:54

0.100 02:40

Test Message: Radio Frequency Interferent Detected

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 605801744 03/30/2017 02:44:18

0.099 02:46

0.097 02:51

Test Message: 0.19

Lab #: 605801745 03/30/2017 04:08:28

0.100 04:10

Test Message: Subject Test Refused

Lab #: 605801746 03/30/2017 04:12:40

0.100 04:14

Test Message: Deficient Sample

Lab #: 605801747 03/30/2017 14:04:09

0.100 14:06

Test Message: Subject Test Refused

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902726 03/05/2017 02:03:37

0.097 02:06

0.097 02:12

Test Message: 0.11

Lab #: 605902727 03/05/2017 02:19:08

0.098 02:21

0.098 02:27

Test Message: 0.10

Lab #: 605902728 03/05/2017 02:37:37

0.099 02:40

0.099 02:45

Test Message: 0.11

Lab #: 605902729 03/05/2017 03:01:52

0.099 03:04

Test Message: Radio Frequency Interferent Detected

Lab #: 605902730 03/05/2017 03:05:17

0.099 03:07

0.097 03:12

Test Message: 0.12

Lab #: 605902731 03/11/2017 00:31:42

0.099 00:34

0.096 00:39

Test Message: 0.17

Lab #: 605902732 03/11/2017 01:03:50

0.098 01:06

0.098 01:12

Test Message: 0.08

Lab #: 605902733 03/11/2017 02:20:40

0.098 02:23

0.097 02:28

Test Message: 0.11

Lab #: 605902734 03/11/2017 02:53:21

0.096 02:56

0.095 03:03

Test Message: 0.08

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902735 03/15/2017 08:35:01

0.098 08:52

Test Message: No Sample Given

Lab #: 605902739 03/18/2017 01:29:38

0.097 01:31

0.097 01:36

Test Message: 0.07

Lab #: 605902740 03/18/2017 02:04:50

0.096 02:07

Test Message: Improper Sample

Lab #: 605902741 03/18/2017 02:13:29

0.097 02:15

0.096 02:20

Test Message: 0.09

Lab #: 605902742 03/18/2017 02:44:16

0.097 02:47

0.097 02:51

Test Message: 0.11

Lab #: 605902743 03/25/2017 23:40:19

0.098 23:44

0.097 23:50

Test Message: 0.19

Lab #: 605902744 03/26/2017 00:14:04

0.097 00:16

0.098 00:21

Test Message: 0.04

Lab #: 605902745 03/26/2017 00:47:51

0.099 00:49

0.097 00:55

Test Message: 0.12

Lab #: 605902746 03/26/2017 01:25:52

0.098 01:28

0.097 01:34

Test Message: 0.07

Lab #: 605902747 03/26/2017 01:45:24

0.098 01:47

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 605902747 03/26/2017 01:45:24

0.097 01:52

Test Message: 0.10

Lab #: 605902748 03/26/2017 02:12:31

0.098 02:14

0.096 02:19

Test Message: 0.16

Lab #: 605902749 03/26/2017 02:21:56

0.097 02:24

0.097 02:29

Test Message: 0.09

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000719

GEBS Results

GEBS VALUE = 0.100

Lab #: 071904250 02/04/2017 02:14:53

0.098 02:17

0.099 02:21

Test Message: 0.09

Lab #: 071904251 02/05/2017 21:26:50

0.099 21:28

0.098 21:33

Test Message: 0.18

Lab #: 071904252 02/05/2017 23:17:44

0.099 23:19

0.098 23:24

Test Message: Return Instrument to Lab

Lab #: 071904253 02/05/2017 23:26:00

0.099 23:27

0.098 23:32

Test Message: 0.18

Lab #: 071904254 02/11/2017 01:54:30

0.098 01:56

0.098 02:01

Test Message: 0.22

Lab #: 071904255 02/12/2017 01:56:10

0.099 01:58

0.100 02:02

Test Message: 0.15

Lab #: 071904256 02/19/2017 01:59:20

0.098 02:01

0.098 02:06

Test Message: 0.16

Lab #: 071904257 02/25/2017 23:08:22

0.100 23:10

0.098 23:15

Test Message: 0.15

Lab #: 071904258 02/28/2017 21:57:30

0.099 21:59

0.098 22:04

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000719

GEBS Results

GEBS VALUE = 0.100

---

Test Message: 0.19



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000726

GEBS Results

GEBS VALUE = 0.100

Lab #: 072600369 02/02/2017 02:22:44

0.100 02:26

Test Message: Improper Sample

Lab #: 072600370 02/02/2017 02:33:47

0.098 02:36

0.098 02:40

Test Message: 0.21

Lab #: 072600371 02/10/2017 00:14:40

0.099 00:16

0.097 00:23

Test Message: 0.16

Lab #: 072600372 02/11/2017 20:33:43

0.100 20:35

0.098 20:43

Test Message: 0.13

Lab #: 072600373 02/12/2017 00:58:58

0.099 01:01

0.099 01:07

Test Message: 0.18

Lab #: 072600374 02/12/2017 02:26:28

0.099 02:29

Test Message: Deficient Sample

Lab #: 072600375 02/12/2017 02:33:50

0.100 02:35

0.098 02:41

Test Message: 0.20

Lab #: 072600376 02/18/2017 02:18:10

0.099 02:20

0.099 02:25

Test Message: 0.15

Lab #: 072600377 02/24/2017 01:14:17

0.100 01:16

0.099 01:22

Test Message: 0.14

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

Lab #: 072703641 02/01/2017 21:09:58

0.097 21:12

0.095 21:17

Test Message: 0.16

Lab #: 072703642 02/03/2017 03:42:36

0.098 03:44

0.096 03:51

Test Message: 0.14

Lab #: 072703643 02/03/2017 22:45:42

0.098 22:48

Test Message: Radio Frequency Interferent Detected

Lab #: 072703644 02/03/2017 22:49:25

0.098 22:52

0.097 22:57

Test Message: 0.12

Lab #: 072703645 02/04/2017 04:02:51

0.098 04:05

0.095 04:11

Test Message: 0.29

Lab #: 072703646 02/04/2017 23:13:34

0.098 23:15

0.093 23:20

Test Message: 0.21

Lab #: 072703647 02/11/2017 03:01:40

0.097 03:04

0.097 03:10

Test Message: 0.08

Lab #: 072703648 02/12/2017 20:48:31

0.097 20:50

0.096 20:58

Test Message: 0.17

Lab #: 072703649 02/18/2017 09:23:29

0.098 09:25

0.096 09:31

Test Message: 0.14

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 072703650 02/20/2017 05:19:01

0.099 05:21

Test Message: Sequence Aborted

Lab #: 072703651 02/24/2017 00:42:35

0.098 00:45

0.097 00:50

Test Message: 0.18

Lab #: 072703652 02/25/2017 04:36:12

0.099 04:39

0.096 04:44

Test Message: 0.12

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000728

GEBS Results

GEBS VALUE = 0.100

Lab #: 072802465 02/01/2017 04:44:41

0.100 04:47

0.099 04:52

Test Message: 0.14

Lab #: 072802466 02/02/2017 02:17:01

0.099 02:19

0.098 02:24

Test Message: 0.22

Lab #: 072802467 02/03/2017 01:41:32

0.099 01:43

Test Message: Deficient Sample

Lab #: 072802468 02/05/2017 22:42:39

0.099 22:44

0.098 22:50

Test Message: 0.15

Lab #: 072802469 02/08/2017 02:12:45

0.099 02:15

0.098 02:20

Test Message: 0.20

Lab #: 072802470 02/10/2017 22:40:43

0.098 22:42

0.098 22:47

Test Message: 0.04

Lab #: 072802471 02/15/2017 01:08:27

0.099 01:10

0.097 01:17

Test Message: 0.12

Lab #: 072802472 02/16/2017 23:09:49

0.099 23:12

0.098 23:17

Test Message: 0.16

Lab #: 072802473 02/17/2017 01:00:29

0.099 01:03

0.098 01:08

Test Message: 0.13

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000728

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 072802474 02/22/2017 00:43:00

0.099 00:45

0.098 00:52

Test Message: 0.07

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500763 02/05/2017 03:49:11

0.100 03:51

0.099 03:57

Test Message: 0.10

Lab #: 605500764 02/24/2017 23:52:25

0.100 23:57

Test Message: Deficient Sample

Lab #: 605500765 02/25/2017 00:02:00

0.099 00:03

0.096 00:10

Test Message: 0.12

Lab #: 605500766 02/25/2017 01:14:15

0.099 01:16

0.099 01:21

Test Message: 0.09

Lab #: 605500767 02/25/2017 01:34:33

0.099 01:36

0.098 01:43

Test Message: 0.09

Lab #: 605500768 02/25/2017 02:43:20

0.098 02:45

Test Message: Sequence Aborted

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801569 02/01/2017 05:26:23

0.099 05:29

Test Message: Deficient Sample

Lab #: 605801570 02/01/2017 05:33:52

0.099 05:36

0.096 05:43

Test Message: 0.14

Lab #: 605801571 02/01/2017 21:01:30

0.098 21:03

Test Message: Deficient Sample

Lab #: 605801572 02/01/2017 21:08:10

0.097 21:10

0.096 21:16

Test Message: 0.22

Lab #: 605801573 02/01/2017 21:19:14

0.096 21:21

0.098 21:26

Test Message: 0.21

Lab #: 605801574 02/01/2017 22:08:15

0.099 22:10

Test Message: Sequence Aborted

Lab #: 605801575 02/02/2017 02:09:44

0.099 02:12

0.098 02:18

Test Message: 0.10

Lab #: 605801576 02/02/2017 02:32:02

0.099 02:34

Test Message: Deficient Sample

Lab #: 605801577 02/02/2017 13:52:42

0.099 13:54

Test Message: Deficient Sample

Lab #: 605801578 02/02/2017 13:59:26

0.099 14:01

Test Message: No Sample Given

Lab #: 605801579 02/03/2017 00:13:37

0.098 00:15

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Test Message: Radio Frequency Interferent Detected

Lab #: 605801580 02/03/2017 00:20:23

0.100 00:22

0.097 00:27

Test Message: 0.21

Lab #: 605801581 02/03/2017 22:29:15

0.099 22:32

Test Message: Deficient Sample

Lab #: 605801582 02/03/2017 22:36:45

0.098 22:39

Test Message: Radio Frequency Interferent Detected

Lab #: 605801583 02/04/2017 02:06:04

0.099 02:07

0.096 02:14

Test Message: 0.17

Lab #: 605801584 02/04/2017 03:50:57

0.098 03:53

0.097 03:58

Test Message: 0.13

Lab #: 605801585 02/04/2017 04:01:01

0.099 04:03

Test Message: No Sample Given

Lab #: 605801586 02/04/2017 23:05:35

0.099 23:08

0.098 23:13

Test Message: 0.10

Lab #: 605801587 02/05/2017 00:26:47

0.099 00:29

0.098 00:34

Test Message: 0.09

Lab #: 605801588 02/05/2017 02:28:07

0.099 02:30

0.099 02:34

Test Message: 0.03

Lab #: 605801589 02/06/2017 01:26:00

0.098 01:27



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801589 02/06/2017 01:26:00

0.097 01:32

Test Message: 0.31

Lab #: 605801590 02/06/2017 19:13:39

0.098 19:15

0.098 19:20

Test Message: 0.24

Lab #: 605801591 02/09/2017 02:53:16

0.098 02:55

0.097 03:01

Test Message: 0.07

Lab #: 605801592 02/09/2017 23:16:27

0.099 23:21

Test Message: Subject Test Refused

Lab #: 605801593 02/09/2017 23:22:49

0.099 23:25

Test Message: Subject Test Refused

Lab #: 605801594 02/09/2017 23:45:32

0.098 23:47

0.098 23:53

Test Message: 0.02

Lab #: 605801595 02/10/2017 01:30:21

0.098 01:32

0.095 01:37

Test Message: 0.17

Lab #: 605801596 02/11/2017 02:58:14

0.098 03:00

Test Message: Radio Frequency Interferent Detected

Lab #: 605801597 02/11/2017 03:01:29

0.099 03:03

0.095 03:08

Test Message: 0.28

Lab #: 605801598 02/11/2017 03:31:59

0.098 03:34

Test Message: Deficient Sample

Lab #: 605801599 02/11/2017 03:38:42

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801599 02/11/2017 03:38:42

0.098 03:40

0.096 03:47

Test Message: 0.18

Lab #: 605801600 02/11/2017 06:41:56

0.099 06:43

0.096 06:48

Test Message: 0.10

Lab #: 605801601 02/11/2017 10:36:34

0.098 10:39

0.096 10:45

Test Message: 0.17

Lab #: 605801602 02/11/2017 23:32:31

0.099 23:34

0.096 23:42

Test Message: 0.22

Lab #: 605801603 02/12/2017 20:36:53

0.098 20:39

0.096 20:44

Test Message: 0.09

Lab #: 605801604 02/13/2017 10:59:56

0.099 11:02

0.097 11:07

Test Message: 0.24

Lab #: 605801605 02/14/2017 00:48:17

0.099 00:50

0.097 00:55

Test Message: 0.17

Lab #: 605801606 02/14/2017 03:16:27

0.098 03:18

0.096 03:24

Test Message: 0.13

Lab #: 605801607 02/14/2017 19:47:20

0.099 19:49

0.096 19:55

Test Message: 0.20

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801608 02/15/2017 01:02:51

0.099 01:20

0.097 01:25

Test Message: 0.11

Lab #: 605801609 02/16/2017 18:35:47

0.099 18:38

Test Message: Subject Test Refused

Lab #: 605801610 02/16/2017 18:45:04

0.098 18:46

0.094 18:53

Test Message: 0.34

Lab #: 605801611 02/16/2017 21:03:56

0.100 21:06

0.098 21:11

Test Message: 0.05

Lab #: 605801612 02/17/2017 10:56:55

0.098 10:59

0.097 11:04

Test Message: 0.16

Lab #: 605801613 02/17/2017 22:27:59

0.099 22:29

0.098 22:36

Test Message: 0.00

Lab #: 605801615 02/18/2017 04:39:55

0.099 04:41

0.098 04:46

Test Message: 0.16

Lab #: 605801616 02/18/2017 15:11:04

0.099 15:14

0.097 15:19

Test Message: 0.11

Lab #: 605801617 02/18/2017 18:27:54

0.099 18:31

0.099 18:37

Test Message: 0.00

Lab #: 605801618 02/19/2017 02:18:27

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801618 02/19/2017 02:18:27

0.099 02:20

Test Message: Sequence Aborted

Lab #: 605801619 02/19/2017 03:29:36

0.099 03:32

0.098 03:37

Test Message: 0.09

Lab #: 605801620 02/19/2017 03:48:42

0.099 03:50

0.097 03:56

Test Message: 0.15

Lab #: 605801621 02/20/2017 05:12:29

0.099 05:14

Test Message: Sequence Aborted

Lab #: 605801623 02/20/2017 19:43:59

0.100 19:45

Test Message: Sequence Aborted

Lab #: 605801624 02/20/2017 20:07:32

0.100 20:09

0.099 20:16

Test Message: 0.07

Lab #: 605801625 02/20/2017 21:53:34

0.100 21:56

0.097 22:03

Test Message: 0.18

Lab #: 605801626 02/21/2017 03:18:37

0.098 03:20

0.099 03:25

Test Message: 0.29

Lab #: 605801627 02/23/2017 02:38:05

0.098 02:39

Test Message: Deficient Sample

Lab #: 605801628 02/23/2017 02:44:36

0.098 02:46

Test Message: Deficient Sample

Lab #: 605801629 02/23/2017 03:07:01

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801629 02/23/2017 03:07:01

0.099 03:09

Test Message: Radio Frequency Interferent Detected

Lab #: 605801630 02/23/2017 03:11:53

0.098 03:13

0.098 03:19

Test Message: 0.17

Lab #: 605801631 02/23/2017 21:05:04

0.100 21:07

0.098 21:12

Test Message: 0.14

Lab #: 605801632 02/23/2017 22:13:48

0.100 22:15

0.099 22:20

Test Message: 0.12

Lab #: 605801633 02/24/2017 12:40:18

0.100 12:42

Test Message: Deficient Sample

Lab #: 605801634 02/24/2017 12:47:36

0.098 12:50

0.095 12:57

Test Message: 0.20

Lab #: 605801636 02/25/2017 02:17:04

0.100 02:19

0.098 02:24

Test Message: 0.13

Lab #: 605801637 02/25/2017 04:35:11

0.100 04:37

0.097 04:43

Test Message: 0.16

Lab #: 605801638 02/25/2017 17:28:32

0.099 17:31

0.099 17:36

Test Message: 0.00

Lab #: 605801639 02/25/2017 23:57:31

0.099 23:59

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Test Message: Sequence Aborted

Lab #: 605801640 02/26/2017 00:01:09

0.100 00:03

0.096 00:10

Test Message: 0.15

Lab #: 605801641 02/26/2017 03:15:49

0.099 03:17

0.097 03:25

Test Message: 0.15

Lab #: 605801642 02/26/2017 05:17:06

0.100 05:21

0.097 05:26

Test Message: 0.16

Lab #: 605801643 02/26/2017 08:12:20

0.099 08:14

0.098 08:19

Test Message: 0.07

Lab #: 605801644 02/27/2017 02:56:46

0.099 02:59

Test Message: Sequence Aborted

Lab #: 605801645 02/28/2017 04:18:11

0.100 04:20

0.097 04:26

Test Message: 0.16

Lab #: 605801646 02/28/2017 22:09:28

0.100 22:11

0.097 22:16

Test Message: 0.17

Lab #: 605801647 02/28/2017 23:22:33

0.100 23:24

0.097 23:30

Test Message: 0.15

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 605902720 02/25/2017 01:17:31

0.098 01:20

0.098 01:25

Test Message: 0.08

Lab #: 605902721 02/25/2017 01:44:01

0.098 01:46

Test Message: Deficient Sample

Lab #: 605902722 02/25/2017 01:51:01

0.098 01:52

0.097 01:58

Test Message: 0.20

Lab #: 605902723 02/25/2017 02:04:15

0.099 02:06

0.098 02:12

Test Message: 0.12

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000719

GEBS Results

GEBS VALUE = 0.100

Lab #: 071904238 01/02/2017 20:52:27

0.098 20:54

0.099 20:58

Test Message: 0.08

Lab #: 071904239 01/04/2017 02:30:02

0.099 02:32

0.098 02:36

Test Message: 0.20

Lab #: 071904240 01/04/2017 23:46:56

0.099 23:48

0.098 23:57

Test Message: 0.19

Lab #: 071904241 01/10/2017 23:04:51

0.099 23:06

Test Message: Sequence Aborted

Lab #: 071904242 01/13/2017 00:36:28

0.099 00:38

0.099 00:42

Test Message: 0.13

Lab #: 071904243 01/16/2017 21:43:29

0.099 21:45

Test Message: Deficient Sample

Lab #: 071904244 01/16/2017 22:20:48

0.099 22:38

0.099 22:43

Test Message: 0.09

Lab #: 071904245 01/21/2017 01:37:58

0.099 01:39

0.099 01:44

Test Message: 0.11

Lab #: 071904246 01/22/2017 21:25:00

0.099 21:28

0.098 21:33

Test Message: 0.17

Lab #: 071904247 01/27/2017 22:39:22



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000719

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 071904247 01/27/2017 22:39:22

0.100 22:41

0.099 22:46

Test Message: 0.19

Lab #: 071904248 01/28/2017 02:00:31

0.100 02:02

0.099 02:06

Test Message: 0.11

Lab #: 071904249 01/30/2017 03:22:03

0.100 03:23

0.099 03:28

Test Message: 0.23

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000726

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 072600366 01/13/2017 23:20:31

0.099 23:24

0.098 23:30

Test Message: 0.17

Lab #: 072600367 01/19/2017 02:27:46

0.099 02:31

0.099 02:37

Test Message: 0.10

Lab #: 072600368 01/19/2017 22:33:10

0.100 22:35

0.099 22:40

Test Message: 0.13

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000727

GEBS Results

GEBS VALUE = 0.100

Lab #: 072703634 01/08/2017 03:26:07

0.098 03:28

0.097 03:33

Test Message: 0.14

Lab #: 072703635 01/15/2017 03:11:22

0.098 03:14

0.096 03:19

Test Message: 0.15

Lab #: 072703636 01/15/2017 05:06:57

0.098 05:09

0.096 05:13

Test Message: 0.13

Lab #: 072703637 01/15/2017 05:28:44

0.098 05:31

0.097 05:36

Test Message: 0.07

Lab #: 072703638 01/16/2017 23:56:32

0.098 23:58

0.096 00:03

Test Message: 0.11

Lab #: 072703639 01/17/2017 10:22:59

0.098 10:25

0.096 10:30

Test Message: 0.15

Lab #: 072703640 01/22/2017 02:03:46

0.098 02:06

Test Message: Deficient Sample

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-000728

GEBS Results

GEBS VALUE = 0.100

Lab #: 072802457 01/03/2017 01:53:30

0.099 01:55

0.098 02:01

Test Message: 0.17

Lab #: 072802458 01/08/2017 02:07:56

0.099 02:10

0.097 02:18

Test Message: 0.12

Lab #: 072802459 01/18/2017 01:04:01

0.099 01:06

Test Message: Deficient Sample

Lab #: 072802460 01/18/2017 01:13:44

0.099 01:16

Test Message: Deficient Sample

Lab #: 072802461 01/21/2017 04:51:06

0.099 04:53

0.099 04:58

Test Message: 0.21

Lab #: 072802462 01/25/2017 02:35:10

0.099 02:37

0.099 02:42

Test Message: 0.08

Lab #: 072802463 01/26/2017 22:54:37

0.099 22:57

0.098 23:02

Test Message: 0.12

Lab #: 072802464 01/31/2017 22:59:57

0.098 23:02

0.098 23:08

Test Message: 0.07

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

Lab #: 605500752 01/07/2017 01:57:24

0.098 01:59

0.096 02:05

Test Message: 0.17

Lab #: 605500753 01/07/2017 02:09:45

0.098 02:11

0.096 02:19

Test Message: 0.15

Lab #: 605500754 01/07/2017 02:24:33

0.099 02:27

0.096 02:34

Test Message: 0.17

Lab #: 605500755 01/07/2017 02:37:22

0.098 02:40

Test Message: No Sample Given

Lab #: 605500756 01/07/2017 02:57:51

0.098 02:59

0.101 03:05

Test Message: 0.14

Lab #: 605500757 01/07/2017 03:08:30

0.098 03:10

0.098 03:15

Test Message: 0.11

Lab #: 605500758 01/22/2017 00:19:43

0.099 00:22

0.097 00:27

Test Message: 0.10

Lab #: 605500759 01/22/2017 00:54:15

0.099 00:56

0.098 01:01

Test Message: 0.13

Lab #: 605500760 01/22/2017 01:23:00

0.099 01:24

0.099 01:29

Test Message: 0.08

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006055

GEBS Results

GEBS VALUE = 0.100

---

Lab #: 605500761 01/22/2017 01:43:40

0.099 01:45

0.096 01:52

Test Message: 0.28

Lab #: 605500762 01/22/2017 02:29:20

0.099 02:31

0.097 02:37

Test Message: 0.13

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801485 01/01/2017 00:45:35

0.099 00:49

Test Message: Deficient Sample

Lab #: 605801486 01/01/2017 00:54:24

0.098 00:56

0.097 01:03

Test Message: 0.11

Lab #: 605801487 01/01/2017 03:04:57

0.099 03:06

0.099 03:11

Test Message: 0.13

Lab #: 605801488 01/01/2017 16:04:39

0.100 16:06

0.099 16:12

Test Message: 0.09

Lab #: 605801489 01/02/2017 04:02:23

0.099 04:05

Test Message: Deficient Sample

Lab #: 605801490 01/02/2017 04:13:33

0.099 04:15

0.094 04:24

Test Message: 0.12

Lab #: 605801491 01/03/2017 01:48:45

0.100 01:51

0.096 01:57

Test Message: 0.27

Lab #: 605801492 01/04/2017 01:16:15

0.099 01:18

0.097 01:23

Test Message: 0.19

Lab #: 605801493 01/05/2017 08:07:06

0.100 08:09

0.099 08:14

Test Message: 0.05

Lab #: 605801494 01/05/2017 16:22:24

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801494 01/05/2017 16:22:24

0.098 16:25

0.095 16:31

Test Message: 0.24

Lab #: 605801495 01/06/2017 03:39:43

0.098 03:42

0.097 03:46

Test Message: 0.11

Lab #: 605801496 01/06/2017 19:41:31

0.099 19:45

Test Message: Sequence Aborted

Lab #: 605801497 01/07/2017 00:04:48

0.099 00:07

0.098 00:13

Test Message: 0.26

Lab #: 605801498 01/07/2017 20:27:12

0.100 20:29

0.100 20:34

Test Message: 0.15

Lab #: 605801499 01/08/2017 00:01:49

0.100 00:03

0.097 00:10

Test Message: 0.11

Lab #: 605801500 01/08/2017 01:53:22

0.100 01:56

0.099 02:01

Test Message: 0.05

Lab #: 605801501 01/08/2017 03:23:20

0.100 03:25

0.098 03:32

Test Message: 0.16

Lab #: 605801502 01/08/2017 17:20:09

0.099 17:22

0.098 17:27

Test Message: 0.25

Lab #: 605801503 01/10/2017 02:55:03



# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801503 01/10/2017 02:55:03

0.100 02:56

0.096 03:03

Test Message: 0.27

Lab #: 605801504 01/11/2017 13:53:32

0.100 13:56

Test Message: Deficient Sample

Lab #: 605801505 01/11/2017 14:04:37

0.099 14:07

Test Message: Subject Test Refused

Lab #: 605801506 01/11/2017 14:17:21

0.098 14:19

Test Message: Deficient Sample

Lab #: 605801507 01/11/2017 14:30:44

0.099 14:33

Test Message: Deficient Sample

Lab #: 605801508 01/11/2017 14:41:03

0.099 14:43

Test Message: Deficient Sample

Lab #: 605801509 01/11/2017 14:48:52

0.098 14:51

0.097 14:56

Test Message: 0.21

Lab #: 605801510 01/12/2017 19:56:22

0.099 19:58

0.097 20:03

Test Message: 0.17

Lab #: 605801511 01/13/2017 00:11:28

0.100 00:14

0.098 00:19

Test Message: 0.18

Lab #: 605801512 01/13/2017 02:17:28

0.099 02:19

0.098 02:26

Test Message: 0.14

Lab #: 605801513 01/13/2017 03:22:38

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801513 01/13/2017 03:22:38

0.100 03:24

0.098 03:30

Test Message: 0.13

Lab #: 605801514 01/13/2017 18:36:59

0.100 18:38

0.100 18:43

Test Message: 0.00

Lab #: 605801515 01/13/2017 23:54:34

0.099 23:56

0.098 00:01

Test Message: 0.09

Lab #: 605801516 01/14/2017 02:02:54

0.099 02:04

0.098 02:09

Test Message: 0.06

Lab #: 605801517 01/14/2017 02:27:53

0.099 02:33

0.098 02:38

Test Message: 0.15

Lab #: 605801518 01/14/2017 09:08:14

0.100 09:15

0.099 09:20

Test Message: 0.02

Lab #: 605801519 01/14/2017 23:52:56

0.101 23:55

Test Message: No Sample Given

Lab #: 605801520 01/15/2017 00:00:45

0.101 00:02

0.097 00:08

Test Message: 0.23

Lab #: 605801521 01/15/2017 03:09:16

0.099 03:11

0.096 03:16

Test Message: 0.22

Lab #: 605801522 01/15/2017 03:51:48

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801522 01/15/2017 03:51:48

0.099 03:54

0.098 03:59

Test Message: 0.20

Lab #: 605801523 01/15/2017 05:33:38

0.100 05:36

0.099 05:41

Test Message: 0.08

Lab #: 605801524 01/17/2017 00:46:08

0.099 00:53

0.097 00:58

Test Message: 0.22

Lab #: 605801525 01/17/2017 20:33:04

0.100 20:35

0.099 20:40

Test Message: 0.07

Lab #: 605801526 01/18/2017 05:19:18

0.100 05:37

0.098 05:42

Test Message: 0.11

Lab #: 605801527 01/19/2017 21:54:06

0.100 21:56

0.097 22:01

Test Message: 0.21

Lab #: 605801528 01/19/2017 23:51:22

0.100 23:53

Test Message: Deficient Sample

Lab #: 605801529 01/19/2017 23:58:06

0.098 00:00

0.098 00:05

Test Message: 0.15

Lab #: 605801530 01/20/2017 03:59:40

0.099 04:03

0.098 04:08

Test Message: 0.08

Lab #: 605801531 01/20/2017 18:58:51

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801531 01/20/2017 18:58:51

0.099 19:01

0.096 19:08

Test Message: 0.21

Lab #: 605801532 01/21/2017 01:57:20

0.100 02:00

0.099 02:05

Test Message: 0.10

Lab #: 605801533 01/21/2017 02:36:36

0.100 02:38

0.098 02:44

Test Message: 0.20

Lab #: 605801534 01/21/2017 10:27:25

0.100 10:29

0.098 10:34

Test Message: 0.23

Lab #: 605801535 01/21/2017 20:07:31

0.100 20:10

0.099 20:15

Test Message: 0.13

Lab #: 605801536 01/21/2017 22:41:15

0.099 22:45

0.098 22:50

Test Message: 0.22

Lab #: 605801537 01/22/2017 01:11:26

0.100 01:13

0.098 01:19

Test Message: 0.17

Lab #: 605801538 01/22/2017 01:28:04

0.100 01:30

Test Message: Sequence Aborted

Lab #: 605801539 01/22/2017 01:35:04

0.100 01:36

0.095 01:42

Test Message: 0.19

Lab #: 605801540 01/22/2017 01:49:33

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801540 01/22/2017 01:49:33

0.099 01:52

Test Message: Sequence Aborted

Lab #: 605801541 01/22/2017 01:56:42

0.099 01:59

Test Message: Sequence Aborted

Lab #: 605801542 01/22/2017 02:15:42

0.100 02:17

Test Message: Deficient Sample

Lab #: 605801543 01/22/2017 02:55:27

0.099 02:57

Test Message: Radio Frequency Interferent Detected

Lab #: 605801544 01/22/2017 03:02:57

0.098 03:04

Test Message: Subject Test Refused

Lab #: 605801545 01/22/2017 03:10:40

0.099 03:12

Test Message: Radio Frequency Interferent Detected

Lab #: 605801546 01/22/2017 03:21:03

0.099 03:23

0.098 03:28

Test Message: 0.17

Lab #: 605801547 01/22/2017 04:21:55

0.099 04:24

0.098 04:32

Test Message: 0.21

Lab #: 605801548 01/23/2017 22:28:19

0.100 22:30

0.099 22:35

Test Message: 0.17

Lab #: 605801549 01/24/2017 23:35:59

0.099 23:38

Test Message: Sequence Aborted

Lab #: 605801550 01/24/2017 23:54:39

0.099 00:12

0.099 00:16

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Test Message: 0.11

Lab #: 605801551 01/25/2017 02:34:57

0.100 02:38

0.098 02:43

Test Message: 0.14

Lab #: 605801552 01/25/2017 03:26:13

0.100 03:28

0.099 03:33

Test Message: 0.22

Lab #: 605801553 01/26/2017 01:44:10

0.099 01:46

0.099 01:52

Test Message: 0.17

Lab #: 605801554 01/26/2017 01:57:32

0.100 01:59

0.098 02:07

Test Message: 0.18

Lab #: 605801555 01/26/2017 22:44:16

0.101 22:47

Test Message: Sequence Aborted

Lab #: 605801556 01/26/2017 23:02:02

0.101 23:04

0.098 23:10

Test Message: 0.16

Lab #: 605801557 01/27/2017 03:22:24

0.099 03:24

0.099 03:30

Test Message: 0.20

Lab #: 605801558 01/27/2017 13:43:55

0.100 13:46

0.098 13:55

Test Message: 0.37

Lab #: 605801559 01/27/2017 17:01:54

0.100 17:04

0.099 17:09

Test Message: 0.19

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006058

GEBS Results

GEBS VALUE = 0.100

Lab #: 605801560 01/27/2017 21:14:55

0.100 21:18

0.098 21:25

Test Message: 0.18

Lab #: 605801561 01/27/2017 23:51:58

0.100 23:54

0.100 23:59

Test Message: 0.08

Lab #: 605801562 01/28/2017 01:58:27

0.101 02:00

0.098 02:06

Test Message: 0.19

Lab #: 605801563 01/28/2017 15:39:31

0.100 15:42

0.098 15:47

Test Message: 0.32

Lab #: 605801564 01/28/2017 21:52:34

0.100 21:54

0.096 22:00

Test Message: 0.19

Lab #: 605801565 01/29/2017 03:19:08

0.101 03:21

Test Message: Subject Test Refused

Lab #: 605801566 01/29/2017 18:32:41

0.100 18:35

0.098 18:41

Test Message: 0.17

Lab #: 605801567 01/31/2017 19:24:28

0.098 19:26

0.098 19:31

Test Message: 0.20

Lab #: 605801568 01/31/2017 22:57:34

0.099 23:00

0.097 23:06

Test Message: 0.12

# SAN DIEGO PD BREATH TEST CALIBRATION CHECKS

INSTRUMENT: 80-006059

GEBS Results

GEBS VALUE = 0.100

Lab #: 605902710 01/07/2017 02:01:08

0.097 02:03

0.097 02:09

Test Message: 0.10

Lab #: 605902711 01/07/2017 02:13:41

0.098 02:16

0.096 02:21

Test Message: 0.16

Lab #: 605902712 01/07/2017 02:24:37

0.098 02:27

0.097 02:32

Test Message: 0.08

Lab #: 605902713 01/07/2017 02:41:05

0.097 02:43

Test Message: Deficient Sample

Lab #: 605902714 01/07/2017 02:48:05

0.097 02:50

Test Message: Deficient Sample

Lab #: 605902715 01/07/2017 03:01:18

0.098 03:03

0.097 03:09

Test Message: 0.08

Lab #: 605902716 01/22/2017 00:17:03

0.098 00:19

0.096 00:24

Test Message: 0.14

Lab #: 605902717 01/22/2017 00:42:51

0.097 00:46

Test Message: Sequence Aborted

Lab #: 605902718 01/22/2017 01:14:22

0.098 01:16

Test Message: Radio Frequency Interferent Detected

Lab #: 605902719 01/22/2017 01:17:54

0.097 01:22

Test Message: Radio Frequency Interferent Detected